

OFFICE OF THE CHIEF DISTRICT MEDICAL & PUBLIC HEALTH OFFICER -cum- DISTRICT MISSION  
DIRECTOR, KENDRAPARA  
[HEALTH & F.W. DEPARTMENT, GOVT. OF ODISHA]

TENDER CALL Reference No. CDM&PHO/RKSDHHKDP/2021-22/PR/01, Dated: 28.12.2021

No 16025 dt 28.12.21

TENDER DOCUMENT FOR SUPPLY OF DRUG, MEDICAL CONSUMABLES, LABORATORY  
CHEMICALS & REAGENTS, X-RAY ARTICLES, INSTRUMENT/ EQUIPMENTS, AYURVEDA  
PANCHAKARMA, BCL ITEMS & WASTE MANAGEMENT ARTICLES ON RATE CONTRACT BASIS  
FOR A PERIOD OF ONE YEAR FROM THE DATE OF APPROVAL OF TENDER

Tel: 06727-233062, 9439994992

LAST DATE & TIME FOR RECEIPT OF TENDER : 27/01/2022 up to 5.00 PM

DATE & TIME OF OPENING OF TENDER  
(COVER-A & COVER-B) : 28/01/2022 at 11.00 AM

PLACE OF OPENING OF TENDER: Conference Hall of the Chief District Medical & Public  
Health Officer, Kendrapara, At/Po/Dist-Kendrapara-  
754211

ADDRESS FOR COMMUNICATION: O/o- The Chief District Medical & Public Health Officer,  
Kendrapara

At/Po/Dist- Kendrapara-754211

E-mail: admomedicalkendrapara@gmail.com

Telephone: 06727-233062/ 9439994992

## TENDER CALL NOTICE

Sealed Quotations are invited by the undersigned from the intending Manufacturers / Suppliers/ Authorized distributors for supply of the items as per list enclosed.


The bidder should submit documents under two bid system i.e. two separate sealed packets one for Technical bid & another for financial bid superscripted on the cover separately & both the bids are to be kept in a single cover superscripted as "TENDER FOR SUPPLY OF MEDICINES /M.CONSUMABLES/ LAB REAGENTS / X-RAY ARTICLES/INSTRUMENT /EQUIPMENT, AYURVEDA PANCHAKARMA, BCL ITEMS & WASTE MANAGEMENT MATERIALS FOR THE YEAR 2021-22"

The Bidders may download the Tender Documents directly from the website available at <http://Kendrapara.nic.in>. The said quotation should reach in the office of the undersigned by Dt 27.01.2022 up to 5.00 PM through Regd. Post / Speed Post only. The tender will be opened on Dt 28.01.2022 at 11.00 AM in the DTU, Kendrapara. The bidders or their authorized representatives may remain present at the time of opening of tender document, the tender document must contain the bidder's full address with Telephone No., Mobile No. and e-mail id. The undersigned is not responsible for any postal delay and reserves all the rights to reject any or all of the tenders without assigning any reason there off.

a) Price of Tender document

Rs. 1000/- (Rupees one thousand) only (Non-Refundable) in shape of DD only in favor of the RKS, DHH, Kendrapara, Payable at Kendrapara (SBI, MAIN BRANCH)

The tender paper will be rejected if the bidder changes any clause or Annexure of the bid.

  
28.12.2021  
Chief District Medical & Public Health Officer-cum-  
District Mission Director,  
Kendrapara

**TERMS AND CONDITIONS FOR ACCEPTANCE OF TENDER FOR SUPPLY OF DRUG, MEDICAL CONSUMABLES, LABORATORY CHEMICALS & REAGENTS, X-RAY ARTICLES, INSTRUMENT AND EQUIPMENTS, AYURVEDA PANCHAKARMA, BCL ITEMS & WASTE MANAGEMENT MATERIALS TO THE CDM & PHO, KENDRAPARA ON RATE CONTRACT FOR A PERIOD OF ONE YEAR FROM THE DATE OF APPROVAL.**

---

1.1 Sealed tenders will be received by Dated 27 / 01 / 2022 up to 5 PM in the office of the Chief District Medical Officer, Kendrapara for purchase of Drug, Medical Consumables, Laboratory Chemicals & Reagents, X-Ray Articles, Instrument And Equipments, AYURVEDA PANCHAKARMA ,BCL ITEMS & Waste Management Materials. Any tender received after the due date & time will be rejected / returned unopened to the sender. The Tenders will be received through Regd. Post /Speed Post only.

1.2 The sealed tenders 'Cover A' (Technical Bid) & Cover-B (Financial Bid) submitted by the bidder will be opened by the Purchase Committee at the Conference Hall of the CDM & PHO Kendrapara on Dt. 28 / 01 / 2022, 11 AM. The bidders or their duly authorized representatives are allowed to be present during the opening of the tenders.

**ELIGIBILITY CRITERIA:**

2.1 Bidder shall be a manufacturer / Distributor / Agency having valid up-to-date manufacturing License / Drug License, Valid ISO Certificate, Good Manufacturing Practices (GMP) Certificate or WHO GMP certificate (in case of drugs only), & having PAN, Last 3 years IT return, Sale Tax / GST clearance certificate (valid up to 30.09.2021)/ GST Registration for all bidders.

2.2 The annual average turnover of the bidder should be Rs 1 Crore or more during the last three financial years certified by a Chartered Accountant 2018-19, 2019-2020, 2020-2021.

2.3 Bidder / manufacturing unit which has been blacklisted for any item either by the Tender inviting authority or by any State Govt. or Central Govt. organization cannot participate in the Tender for that item during the period of blacklisting supporting to court affidavit (Annexure - IV).

2.4 The Bidder should have typically not less than 3 (three) years of marketing experience in lab. Materials, Medicines / pharmaceutical products / Instruments supplying to the Government organization/ Corporate Hospital / PSU Hospitals/ UN agencies in India duly supported by documentary evidence (Xerox copy of Purchase Order minimum five).

**TECHNICAL BID (COVER - A):**

3.1. A Non-refundable A/c Payee Demand Draft of Rs.1000/- (Rupee one thousand only) towards the cost of the tender document in favor of the RKS, DHH, Kendrapara, Payable at Kendrapara (SBI, MAIN BRANCH)

- 3.2. 20,000/- as EMD in shape of Demand Draft in favor of the RKS, DHH, Kendrapara, Payable at Kendrapara (SBI, MAIN BRANCH)
- 3.3. List of items quoted with strength / specification and packing. Local Medium & Small Enterprises (MSEs) registered in Odisha under the respective DICs, Khadi, Cottage & Handicrafts Industries, OSIC, NSIC are **exempted** from submission of EMD)
- 3.4. Valid up-to-date Manufacturing License / Drug License (In case of Drugs only) up to march 2022.
- 3.5. Detail name, address, telephone no., Mobile No., Fax, e-mail of the firm and of the Director / Managing Partner / Proprietor of the firm. (As per Annexure - II).
- 3.6. Address, Telephone No., Mobile No., e-mail, Fax of the Branch Office / Contact Person in Odisha (Annexure-IV).
- 3.7. The Bidder has to submit their valid Sale Tax/ GST Clearance Certificate (valid up to 30.09.2021), PAN Card & Last 3 years I.T return i.e. 2018-19, 2019-20, 2020-21.
- 3.8. Valid up-to-date WHO GMP/Good Manufacturing Practice (GMP) certificate of Manufacturer from the Competent Authority (In case of Drugs only).
- 3.9. Valid and up-to-date ISO Certificate of the Manufacturer.
- 3.10. The original tender document duly sealed & signed in each page.
- 3.11. Checklist with detail of the documents enclosed in Cover 'A' (as per Annexure - I). The documents should be serially arranged as per this Annexure V and should be securely tied and bound.
- 3.12. If the Bidder is an authorized distributor, the purchase order may be placed to the authorized distributors & bills can be raised by the authorized distributor on behalf of the manufacturer, if specifically authorized by the manufacturer.

**COVER - B (PRICE BID):**

- 4.1. The tender form giving the rates for various Lab. Materials, Consumables, Reagents, Medicines and Medical Consumables etc. should be submitted in a separate sealed cover hereafter called **Cover 'B' (Price Bid)**.
- 4.2. The tender format (price schedule) in the prescribed form (**as per Annexure - II**) should be submitted inclusive of all taxes i.e. excise duty, insurance, packing, Central Sales Tax / VAT & Entry Tax Gst etc. forwarding and freight (i.e. door delivery). The rate should be quoted for Unit pack (**both in figures and words**). & supply will be made in unit pack as per Tender

specifications. The hard copy of price schedule must be sealed & signed in each page by the Bidder.

- 4.3. The Cover 'B' of the successful bidders who will qualify in the Technical bid will be opened at the Office chamber of the DTU, Kendrapara by the members of Purchase Committee in the presence of the Bidders or their authorized representatives if they like so.

**NON RESPONSIVE / REJECTION CRITERIA**

5. The tender paper will be rejected if any of the following documents are wanting.
- i) Earnest Money Deposit (EMD) Rs. 20000/- (Rupees twenty thousand) only *Local Medium & Small Enterprises (MSEs) registered in Odisha under the respective DICs, Khadi, Cottage & Handicrafts Industries, OSIC, NSIC are exempted from submission of EMD* in favour of the RKS, DHH, Kendrapara payable at Kendrapara.
  - ii) Bank Draft of Rs.1000/- towards price bidding document in favour of the RKS, DHH, Kendrapara, payable at Kendrapara.
  - iii) Attested photocopies of up-to-date valid manufacturing license / Drug License (in case of Drugs Only).
  - iv) Duly attested valid up-to-date Good Manufacturing Practices (GMP) Certificate or WHO GMP certificate (in case of Drugs Only)
  - v) Valid and up-to-date ISO Certificate of the Manufacturer.
  - vi) Price Bid / quoted rates (hard copy) without signature and seal of Bidder.
  - vii) Attested photo copy of Purchase Order from the Government organization/ Corporate Hospital / PSU Hospitals/ UN agencies in India in support of marketing experience for last 3 years in Tender for supply of Lab. Materials, Consumables, Reagents, Medicines & Medical Consumables (minimum five order).
  - viii) Attested copy of Last 3 years IT return & PAN Card.
  - ix) Attested copy of valid GST clearance certificate, (up to 30.09.2021).
  - x) Authorization letter from the Manufacturer if the Bidder is an authorized distributor.
  - xi) The annual average turnover of the bidder should be Rs 1 Crore or more during the last three financial years certified by a Chartered Accountant.

**EARNEST MONEY DEPOSIT (BID SECURITY):**

- 6.1 The EMD is to be deposited to safeguard against a bidder's withdrawing or altering its bid during the bid validity period. The Earnest Money Deposit Rs. 20,000/- (Rupees twenty thousand only) (Not Applicable for SSI UNITS) for tender will be submitted by the Bidder. The EMD should be paid in the shape of demand draft in favor of the RKS, DHH, Kendrapara from any Nationalized / Scheduled Bank payable at Kendrapara. This should be submitted with the tender in Cover 'A'. E.M.D. in shape of Cheque / Cash / Postal Order will not be accepted.

- 6.2 The E.M.D will be forfeited if the Bidder withdraws the tender / violate the tender condition in any respect within the validity of the bid or does not accept the approved rates by the bidder.
- 6.3 The E.M.D of the unsuccessful bidders only will be returned back without interest latest by the 30<sup>th</sup> days after expiry of validity of the bid.

**TENDER CONDITIONS:**

- 7.1 The quoted rate shall not vary with the quantum of order placed for destination points.
- 7.2 Tenders should be type-written or Computerized and every correction /over-writing in the tender should invariably be attested with signature of the Bidder with date before submission of the tenders to the authorities concerned. No revision of price (upward or downward) will be allowed once the tender is opened ("Cover B").
- 7.3 If there is a discrepancy between words and figures, **the amount mentioned in words will prevail.**
- 7.4 If there is discrepancy in absolute and unit rate (which is obtained by multiplying the absolute rate with the total number in a unit pack) **the absolute price will prevail and the unit price will be corrected accordingly.**
- 7.5 To ensure sustained supply without any interruption the CDM & PHO, Kendrapara reserves the right to split orders for supplying the requirements among more than one Bidder provided that, the rates and other conditions of supply are equal and with sufficient grounds. In case of non-supply of any item by any approved lowest quoted firm, willingness of L<sub>2</sub> / L<sub>3</sub> / L<sub>4</sub> / L<sub>5</sub> etc. will be obtained to supply at L<sub>1</sub> rate. The unit packing should be uniform as prescribed for each item.
- 7.6 **The rates quoted and accepted will be binding on the Bidder for a period of one year from the date of publication of the approved list (Opening date of Price Bid) or that of publication of the next approved list whichever is earlier and on no account any increase in the price will be entertained till the completion of this tender period.**
- 7.7 No Bidder shall be allowed at any time on any ground whatsoever to claim revision of or modification in the rates quoted by him. Clerical error, typographical error etc. committed by the Bidders in the tender forms will not be considered after opening of the tenders. Conditions such as **"SUBJECT TO AVAILABILITY, SUPPLY WILL BE MADE AS AND WHEN SUPPLIES ARE RECEIVED"** etc. will not be considered under any circumstances and the tenders of those who have given such conditions shall be treated as incomplete and for that reason, shall be summarily rejected.
- 7.8 If at any time during the period of contract, the price of tendered items is reduced or brought down by any law or act of the Central or State Govt. or by the Bidder himself, the

Bidder shall be morally and statutorily bound to inform the CDM & PHO immediately about such reduction in the contracted prices. The CDM & PHO Kendrapara is empowered to unilaterally effect such reduction as is necessary in rates in case the Bidder fails to notify or fails to agree for such reduction of rates. In case of any enhancement in Excise Duty due to statutory Act of the Govt. after the date of submission of the tenders and during the tender period, the additional excise duty so levied will be allowed to be charged extra as separate item without any change in price structure of the drugs approved under the tender. For claiming the additional cost on account of the increase in excise duty, the Bidder should produce letter from the concerned excise authorities indicating his commitment for the supply made to the CDM & PHO Kendrapara on account of the increase in excise duty.

- 7.9 The Tax will be charged as per the guidelines given by the Finance Deptt. From time to time. Only one Sales Tax either C.S.T / VAT will be paid to the supplier. In case of Entry Tax the supplier has to deposit the original receipt to claim it, if finished goods are brought from outside the State. The Sales Tax / VAT & E.T component should be shown separately in the cost price (Price Schedule) (Annexure – II).
- 7.10 In the event of the date being declared as a holiday for Govt. of Odisha, the due date of submission of bids and opening of bids will be the following working day at the appointed place & time.
- 7.11 The authority will be at liberty to terminate by assigning any reason there of the contract either wholly or in part. The Bidder will not be entitled for any compensation whatsoever in respect of such termination.
- 7.12 If the Bidder fails to execute the supply within the stipulated time, the authority is empowered to make emergency purchases or purchase from L<sub>2</sub>, L<sub>3</sub>, L<sub>4</sub> & L<sub>5</sub> as the case may be and apply the penal clauses on the defaulting firm.
- 7.13 The quantity of procurement may increase or decrease depending on the budget provisions, new schemes, natural calamities and epidemics etc.
- 7.14 The Bidder has to accept all the term & conditions mentioned in the tender document.
- 7.15 **The tender document is not transferable.**

#### **SUPPLY CONDITIONS:**

- 8.1 The supplier / firm will supply as per the technical specification and print the generic name in bold letters and brand name if approved by the licensing authority may be printed in small font in strip / bottle / injection / packing / foil / carton / packet / box etc.
- 8.2 The Bidder should deliver the Lab. Materials, Consumables, Reagents, Medicines and Medical Consumables at the Consignee (Door Delivery) in the district. The insurance, storage & transport charges / courier charges if any will be borne by the supplier. The short

- supply, damage if any at the time of delivery of consignment shall be replaced by the supplier within seven days of the first supply of indented items.
- 8.3 In all the cases the responsibility of the purchaser will start only after delivery and due verification of goods.
- 8.4 The Composition & strength of each item tendered should be as per the specification given in (Technical Specification).
- 8.5 The supply should be started immediately & completed within 60 days from the date of issue of the purchase order. If supply is not completed within 60 days from the date of issue of purchase order, the authority may cancel the order or allow extension. The authority has also the liberty to cancel those orders and purchase the same item from L<sub>2</sub>, L<sub>3</sub>, L<sub>4</sub> & L<sub>5</sub> firm as the case may be if the other firms agree to supply at L<sub>1</sub> rate. The non-supplier will be penalized in form of forfeiture of E.M.D.
- 8.6 Each Drugs and Medical Consumables must be accompanied with a test certificate that the supplied Drugs and Medical Consumables is of standard quality and as per the standard.
- 8.7 The Lab Materials, Drugs and Medical Consumables should arrive at the destination (warehouse) with remaining shelf-life of at least 5/6<sup>th</sup> of the stipulated total shelf-life from the date of manufacturing of that product.
- 8.8 The labels in all case of injectables should clearly indicate whether the preparations are meant for INTRA VENOUS, INTRA-MUSCULAR or SUB-CUTANEOUS or INTRADERMAL etc.
- 8.9 The packaging cartons must bear the name of the items (**Generic names**), strength, total quantity, total weight, and name of the manufacturer, month of manufacturing and month of expiry. All the supplied medicine should be super scribed as "ODISHA GOVT SUPPLY, NOT FOR SALE"
- 8.10 Each Strip / Box / Carton / Bottle / Amp. / Vial/ Packet shall bear the seal of the manufacturer and month of manufacturing, month of expiry & Batch No/Lot No.

**QUALITY TESTING:**

- 9.1 All the suppliers / manufacturers must submit test reports from approved laboratory of the batches of Drugs Medicines and Medical Consumables being supplied to the consignee.
- 9.2 The district will also do random testing of the supplied drugs if required for which quality testing, packing & forwarding charges if any will be borne by the supplier.
- 9.3 If the Drugs and Medical Consumables as per report is found not of standard quality in first test, the supplier will be required to replace the entire quantity of the batch declared NSQ or the cost of it in shape of demand draft in favour of the RKS, DHH, Kendrapara & take back the available stock (unused) in different health institutions of Kendrapara district in his own cost within a period of 30 days of receipt of letter from authority.

- 9.4. If the supplier does not take back the NSQ stocks or does not challenge the 1<sup>st</sup> test report within 30 days of issue of letter by the authority, the concerned health institution can destroy the NSQ stocks with intimation to the authority following criteria mentioned in the Notification no.25061/H dt.22.09.2012 of Health & FW dept. Govt. Odisha.
- 9.5. If the items has been tested by the D.I & the CDM & PHO Kendrapara, the NSQ report of either of the authority will be accepted & actions will be taken accordingly. If the supplier challenges the test report in the above circumstances then it will go for testing to Central Drugs Testing Laboratory, whose report will be final.
- 9.6. The supplier shall be responsible for the full replacement in his own cost for any product, if the same is found on visual inspection to have deteriorated / not of standard quality before the expiry date.

**TERMS OF PAYMENT:**

- 10.1. No advance payments towards cost of Lab. Materials, Consumables, Reagents, Medicines and Medical Consumables etc. will be made to the supplier. The supplier has to submit 3 (three) copies of the bills or invoices with a photocopy of the purchase order at the place of supply for stock entry.
- 10.2. No claims shall be made against the authority in respect of interest on earnest money deposit or delayed payment.

**PENALTIES:**

- 11.1. If the Bidder withdraws or alters its bid or unwilling to accept the term & conditions of the tender after submission of bid & during the bid validity period, the EMD deposited by the said bidder will stand forfeited.
- 11.2. If the successful Bidder fails to execute the work order / replacement of NSQ stock within the time specified/ unable to undertake the contract or supply as per purchase order, his purchase order will be cancelled by him shall stand forfeited. He will also be liable for the damages sustained by the authority. Such damages shall be assessed by the authority, whose decision will be final in the matter. The defaulting firm will be de-recognized for 2 (two) years from the date of issue of letter. The authority can forfeit the performance security of the defaulting firms and de-recognize / black list the said firms for 2 (two) years from the date of issue of letter.
- 11.3. If any Drugs and Medical Consumables supplied by the Bidder being partially used or consumed after supply and are subsequently found to be in bad order, unsound, inferior in quality or description or otherwise faulty or unfit for consumption, the supplier will be required to replace the entire quantity of that item or the cost of it in shape of bank draft in favour of the RKS, DHH, Kendrapara and take back the available stock (unused) present in

different health institutions of Kendrapara district at his own cost within a period of 30 days of receipt of letter from CDM & PHO Kendrapara and also the firm will be de-recognized for 2 (two) years for that item from the date of issue of letter.

- 11.4 In the event of any litigation arising out of the tender such matters would be subject to the jurisdiction of the Courts of Kendrapara.

The detail items are required as follows:

1. LIST OF DRUGS, MEDICAL CONSUMABLES.
2. LIST OF MENTALHEALTH DRUGS.
3. LIST OF NCD DRUGS.
4. LIST OF X-RAY & NIDAN LABORATORY CHEMICALS, ANTIGEN, PIPETTE & VIALS & CONSUMABLES.
5. LIST OF BEDDING, CLOTHING, AND LINEN ITEMS.
6. LIST OF OPH LABORATORY ITEMS.
7. LIST OF LABORATORY MATERIALS REQUIRED FOR RNTCP USE.
8. LIST OF ITEMS REQUIRED FOR DEIC UNIT.
9. List of PHYSIOTHERAPY EQUIPMENT TO BE PURCHASED FOR DEIC.
10. List of Equipment's and Furniture and Diagnostic Kits for UPHC
11. LIST OF EQUIPMENT, CONSUMABLES AND MISCELLANEOUS SUPPLIES FOR HWC.
12. LIST OF ITEMS OF PHYSIOTHERAPY & GERIATRIC EQUIPMENT & INSTRUMENT
13. LIST OF WASTE MANAGEMENT ARTICLE
14. LIST OF PANCHAKARMA AYURVEDA MEDICINE

1. LIST OF MEDICINES, MEDICAL CONSUMABLES etc

Sl no	Drug Code	Drug Name	Specification
1		Cap Calcigard	10mg
2		Cap Calcigard SR	20mg
3	D04051	Cap Indomethacin	25mg
4		Cap Multi-vitamin & mineral	
5		Cap Nifedipine	5Mg/
6	D32015	Cap Oseltamivir	75mg
7		Cap Pancreatic Enzymic	10000&30000 USP Unit
8		Cap Rabeprazole + Levosupride	
9		Cap. Itraconazole	200Mg
10	D18038	Cap. Itraconazole	100Mg
11		Cap. Micro Progesteron	100mg/cap
12	D26017	Cap. Micro Progesteron	200mg /cap
13	D21009	Cap. Omeprazole	20Mg
14	D30003	Cap. vitamin B complex	
15	D34001	Cap.Vit. A.D	Vit. A 5000 IU+vit. D3-400 IU/Cap
16		Tab Pyridoxin	100MG
17		Tab Nitroglycerine	
18		Tab Sorbitrate	
19	D12015	Tab .Terbinafine	250Mg/Tab
20	D04025	Tab Aceclofenac	100Mg/Tab
21	LD0082	Tab Amlodipine 5 mg + Atenolol 50 mg	5mg+50mg
22		Tab Aspirine	75mg
23	D27095	Tab BetahestIn	15mg
24		Tab Bisacodyl	5mg
25	D09091	Tab Cefixime	200Mg
26	D09126	Tab Cefodoxime	200mg
27		Tab Cefpodoxime proxetil+Clavonic	200mg+125/tab
28		Tab Cefuroxime	250mg
29	LD0021	Tab cefuroxime Axetil	500mg
30	D05010	Tab Cetrizine	10Mg
31		Tab cholecalciferol	
32		Tab Cinnarizine	25mg/tab
33	D07024	Tab Citicolin	500mg
34		Tab Clofazimin	100mg
35	D27110	Tab Clonazepam	0.5mg/tab
36		Tab Clopidrogel+Aspirine+Atrovastine	150mg+150mg+40mg
37	D17048	Tab Clopidrogel	75mg
38		Tab Deflazacort	6/12/24/30

39	LD0049	Tab Dicyclomine 20mg+PCM 500mg	
40	D27035	Tab Divalproex Sodium	500mg
41	D27033	Tab Dothiepin	75mg
42	D21033	Tab Drotavrerin	80Mg/Tab
43		Tab Ethambutol	800mg
44		Tab Ethambutol	400mg
45		Tab Ethionamide	250mg
46		Tab Fexo Fenadine	120mg
47	D46004	Tab Fleroxate	200mg
48	D12007	Tab Fluconazole	50mg
49	D20001	Tab Frusemide	40mg/t
50		Tab Glycerol trinitrate	10mg
51		Tab Hydrochlorothiazide	12.5mg
52		Tab Isonizide	300mg
53	D17001	Tab Isosorbide	5mg
54		Tab Isosuprine	
55		Tab Lansoprazole	50mg/t
56	D05021	Tab Levocetirizine	5Mg
57		Tab Levofloxacin	500mg
58		Tab Levofloxacin	250mg
59		Tab Linezolid	600mg
60	D17051	Tab Losartan Potassium	50Mg/Tab
61	D26003	Tab Methylergometrine maleate	0.125Mg/Tab
62		Tab Metoprolol	50mg
63	D13005	Tab metronidazole	400Mg
64	D26012	Tab Mifepristone	200mg/Tab.
65		Tab Moxifloxacin	400mg
66	D17052	Tab Nifedipine SR	20Mg/SR Tab
67	D09033	Tab Ofloxacin	200Mg
68		Tab ondansetron	8mg
69	LD0143	Tab Pantoprazole 40mg + Domperdone 10mg-	40mg +10 mg
70	D04002	Tab Paracetamol	500Mg/.Tab
71	LD0095	Tab Paracetamol	650Mg/Tab
72	D05015	Tab prednisolone	10Mg
73		Tab Pyrazinamide	750mg
74		Tab Pyrazinamide	500mg
75		Tab Pyridoxine	100mg
76	D27054	Tab Quetiapine	100mg
77		Tab Rifampicin	150mg
78		Tab Rifampicin	450mg
79		Tab Rifampicin	300mg
80	D17071	Tab Rosuvastatin	10mg

81		Tab Rosuvastine	20mg
82		Tab Salbutamol	4mg
83	D04056	Tab serratiopeptidates	10mg
84		Tab Stemetil	5mg
85	D17084	Tab Telmisartan	20mg
86		Tab Telmisetron+Hydrochloro thiazide	40+12.5mg
87	D28003	Tab Theo+Eto	
88	D20011	Tab Torsiamide	20mg
89	D04030	Tab Tramadol	100mg
90	D30035	Tab Vit D3	60000 IU
91		Tab Voglibose	0.3 mg
92	D09077	Tab, Azithromycin	500Mg
93	LD0050	Tab, Aceclofenac+ Paracetamol	100mg+325mg
94	D09113	Tab, Amoxycilin +Clavulanic Acid	500mg+125mg
95	D30001	Tab, Calcium and Vit, D3 /Tab	(Film coated)(Aluminium foil/Glycine Poly Foil/Blister pack)-500mg Elemental Calcium + Vitamin D3 250IU
96	D09026	Tab, Cefadroxil	500Mg
97	D05006	Tab, CPM	4Mg
98		Tab, DEC	150mg
99	D04023	Tab, Etoricoxib	90mg
100	D12008	Tab, Fluconazole*	150Mg
101	LD0009	Tab, Glimipride + Metformin	1mg+500mg
102	D22004	Tab, Glipizide	5Mg/Tab
103	D05028	Tab, Levocetirizine + Monteleukast	5mg +10mg Tab
104		Tab, MedoxyProgesteron Acetate	10Mg/ Tab
105	D26007	Tab, Misoprostol	200Mg /tab
106	D09084	Tab, Ofloxacin	400Mg
107	LD0084	Tab, Ofloxacin+Orindazole	200mg & 500 mg
108		Tab, Orneprazole+Domperidone	
109	LD0058	Tab, Rabeprazole +Domperidone -	
110	D30016	Tab, Vitamin C	500mg
111	LD0033	Tab,Aceclofenac + Paracetamol + Sera	
112	LD0022	Tab,Cefurexime+ Clavulanic Acid	500mg+125mg
113	D27001	Tab,Diazepam (Aluminium foil/Blister pack)-	5 mg/Tab
114		Tab,Diclofenac Sodium	50Mg/.Tab
115	D04006	Tab,Ibuprofen	400Mg/Tab
116	D16013	Tab,Tranexamic acid	500mg/Tab
117	D21048	Tab/Cap, Rabeprazole+ Domeprodone	20mg+30mg
118		Inj Drotavrerin	
119		Inj Efpress	
120		Inj Medroxyprogesteron	

121	D39002	Inj 8%Amino Acid for liver	
122	D04034	Inj Aceclofenac 150mg	
123		Inj Amino phylline	250mg/2ml
124	D09112	Inj Amoxycillin+ Calvunac Acid	1gm+200mg
125	D09162	Inj Ampicillin +Cloxacilline	250mg+250mg
126	D09047	Inj Ampicillin Sodium	100mg
127		Inj Ampicillin Sodium	250mg
128		Inj Ampicillin Sodium	500mg
129	D23009	Inj Anti Rabies Vaccin for human use.	2.5IU
130	D03001	Inj Atropine	0.6 mg/ml
131		Inj Betamethasone	
132	D04039	Inj Butorphanol Tartrate	2mg/ml
133		Inj Calcium Gluconet	
134		Inj Capreomycin	750mg
135	D09078	Inj Cefoperazone + Salbactam	500mg +125mg
136	D09054	Inj Cefotaxime	1g/ V
137		Inj Cefotaxime	500mg
138	D09042	Inj Cefotaxime	125mg
139		Inj Ceftadigim + Tazobactom	1.125gm
140	D09116	Inj Ceftriaxone +Salbactum	1.5gm
141	D07023	Inj Citicolin	250mg/2ml
142		Inj Clindamycin	150mg
143	D05001	Inj Dexamethasone	4.8 mg
144		Inj dextrose 25%	100 ml bot
145		Inj Dobutamine	
146		Inj Dopamine	
147	D16047	Inj Erythropoietin	4000IU
148		Inj Etophylline+Theophylline	
149		Inj Expovan (Plasma Expander )	500ml
150		Inj Ferric Carboxy maltose	1000mg/vial
151	D20002	Inj Frusemide	10Mg/1Ml
152		Inj Gentamycin	20mg
153		Inj Gentamycin	80mg
154		Inj Hepatitis B	5ml/Vial
155	D40002	Inj Human Albumin IV	20%
156		Inj Isosuprine	10mg
157		Inj Kanamycin	500mg
158		Inj Kanamycin	1000mg
159		Inj KCl	10Ml*
160	D04035	Inj Ketorolac	30mg
161	D02001	Inj Lignocaine HCL	2%w/v
162		Inj Lincormycin	2ml

163		Inj Lopamide	40mg
164		Inj Lorazepam	
165		Inj Mannitol IV	20%100ml
166	D09163	Inj Meropenem	500mg/vial
167		Inj Meropenum	1gm v
168		Inj Methyl cobalamine	
169		Inj metoclopramide	
170		Inj Metronadizole IV	100ml
171		Inj Midazolam	1mg/ml
172		Inj Na Cl	100ml bot
173		Inj Neteromycin	300mg
174	D09157	Inj Netilmycin	10mg/vial
175	D05018	Inj Noradrenaline	1mg/ml
176		Inj NS 3%	100ml/bot
177	D40004	Inj Octreotide	100mg
178		Inj Ofloxacin IV	100ml
179	D04004	Inj Paracetamol	150mg
180	D05007	Inj pheniramine maleate	22.75mg/ml
181	D07002	Inj phenobritone Sodium	200mg/ml
182	D04014	Inj Piroxicam	20mg
183		Inj Prochlorperazine	12.5mg/ml
184		Inj Tramadol	50mg/ml
185	D05005	Inj. Adrenaline Bitartrate-	1mg/ml (1:1000)
186	D09019	Inj. Amikacin	500Mg
187	D09046	Inj. Amikacin	100Mg
188	D09004	Inj. Ampicillin	500Mg
189	D09056	Inj. Ampicillin	250mg
190	LD0006	Inj. Caffeine Citrate	2ml/vial
191		Inj. Ceftriaxone	125mg/v
192	D09094	Inj. Ceftriaxone	1g/V
193	D09093	Inj. Ceftriaxone	250mg
194	D09115	Inj. Ceftriaxone + Tazobactam	1gm + 125mg / vial
195	D07006	Inj. Diazepam-	5mg/ml
196	D23011	Inj. Human Anti-D Immunoglobuline	300mcg/2ml
197	D05002	Inj. Hydrocortisone	100Mg/V
198	D16018	Inj. Iron Sucrose	(2ml ampie)
199	D17043	Inj. Labetalol	20mg/4ml Amp
200	D02005	Inj. Lignocaine with Dextrose (Heavy)	53.3Mg+75Mg/ MI
201	D02003	Inj. Lignocaine+ Adrenalline	21.3+0.005Mg/MI
202	D26006	Inj. Magnesium Sulphate	500Mg/ml
203	D26001	Inj. Methylergometrine maleate	0.2Mg/MI
204	D05023	Inj. Methylprednisolone	40mg/Vial

205	D24006	Inj. Neostigmine Methylsulphate-0.5 mg/ml	1ml
206	D21024	Inj. Ondansetron	2mg/ml(2ml amp)
207	D06002	Inj. Pam	500mg/20ml Vial
208	D21030	Inj. Pantoprazole With Diluent	40Mg/Vial
209	D04008	Inj. Pentazocine Lactate	30Mg/MI
210	D07007	Inj. Phenytoin Sodium-	50 mg/ml
211	D09087	Inj. Piperacilline+Tazobactam	4.5g
212	D23008	Inj. Snake Venom Antiserum	10 ml/Vial
213	D23005	Inj. Tetanus Toxoid	5ml/vial
214	D16012	Inj. Tranexamic Acid	500Mg/5MI
215	D02006	Inj. Bupivacaine(0.5% Heavy)	5Mg/ML
216	D26005	Inj. Carboprost	250mcg/MI
217	D04049	Inj. Diclofenac sodium Aquous	75mg/ml
218	D26002	Inj. Oxytocin	5IU/1MI
219	D21005	Inj. Promethazine HCL	25Mg/
220	D03002	Inj. Vit B Complex	2MI/Amp
221	D09095	Azithromycin Oral Suspension	200mg / 5ml
222		Calcium Vit D3 Susp	
223	LD0056	Cough Expectorant	100ml
224		Drotaverin syp:	
225	D07011	Phenobarbitone Oral Solution	20mg/5ml
226		Sus Norfloxacin	200mg/5ml
227		Susp Cefixime	50mg/5ml
228		Susp Cefuroxime	125mg/5ml
229		Susp Dexamethasone	
230		susp Lincomycin	300mg/ml
231	D09085	Susp Ofloxacin	100mg
232	D21027	Susp. Antacid	100ML Bottle
233	D21021	Susp. Sucralfate	100MI
234		Syp Cefodoxime	50mg
235		Syp Flexo fenadine	60mg
236		Syp Hydroxyzin	
237		Syp Ibu+Paracetamol	
238	D40003	Syp Lactulose	
239		Syp Laxative	200ml/Bottle
240		Syp Levo salbutamol +Guaphenesin +Ambroxol	1MG+30MG+50MG
241		Syp Levofloxacin	125mg/5ml
242		Syp Mefenamic+Paracetamol	
243	LD0140	Syp Multivitamin-	200ml/Bottle
244	D21006	Syp promethazine	5 mg/5 ml
245		Syp Salbutamol+Ambroxil	1mg+15mg/5ml

246		Syp Terbutalin + Ambroxil	1.25MG+50MG
247		Syp vitamin B complex	
248		Syp. Aktlyser	
249	D09164	Syp. Amoxycillin + Clavunic Acid	200 mg + (28.5 mg)
250	D05011	Syp. Cetirizine	5Mg/5MI
251	D05016	Syp. Levocetirizine	2.5mg/5ml
252		Syp. Levocetirizine +Montelukast	2.5mg+5mg/5ml
253	D21038	Syp. Ondansetron	2 mg/5ml
254	D04003	Syp. Paracetamol	125Mg/5MI
255	D30024	Drop Multivitamin	(Palatable, with dropper and container as per I.P).
256		Acyclovir Eye Oint.	0.3%w/v
257	D09109	Amoxycillin Dry Syrup	
258	D25046	Carboxy Methyl Cellulose Eye Drop	0.50%
259		Chloromycetin ear drop	5mg/ml
260		Cyclopintolate Eye Drop	1%/10mg
261		Dexamethasone +Moxifloxacin	0.1mg+0.5mg/5ml
262		Drop Ampicillin	100mg/ml
263	D09105	Drop Cefixime	25mg/ml
264		Drop Dicyclomine	
265		Drop Drotaverin	
266		Drop Lactage	
267		Drop LevoSalbutamol + Ambroxol + Guaiphenesin	0.25mg+7.5mg+12.5mg/ml
268		Drop Ondansetron	
269	D04026	Drop Paracetamol	100 mg/ 1 ml
270		Drop. Amoxycillin + Clavunic Acid	80mg+11.4mg
271		Drop.Cefodoxime	25mg/ml
272		Drop.Vitamin D3	400IU/MI,30ML
273		Drop Amoxycillin	125mg/ 5ml
274	D25026	Lidocaine+ Ofloxacin Ear Drop	5MI / Vial
275		Lotiprodml Eye Drop	
276		Lubricating Eye Drop	
277		Midazolam Nasal Spray	
278	D25034	Moxifloxacin Eye Drop	0.5% (FFS/
279		Moxifloxacin+Dexamethasone Eye Drop	0.5%+0.1/5ml/vial
280		Natamycin Eye Drop	
281	D25041	Ofloxacin Demamethasone Eye/ear Drop	
282	D25025	Ofloxacin Eye Drop	0.3%w/v
283		Ofloxacin eye/ear drop	0.5ml
284		Clopatadine Eye Drop	
285		Oxymetazoline Nasal Drop Adult	0.05%

286		Oxymetazoline Nasal Drop Ped	0.03%
287		Procaline Eye Drop	
288	D29015	Saline Nasal Drop	10MI Vial
289		Timolit Eye Drop	
290		Tropicamide Plus Eye Drop	0.5%/5ml
291		Xylocaine 4% Topical Eye Drop	4%
292	D25045	Xylometazoline Hydrochloride Nasal Drop- (Adult)	0.1 wv FFS BFS Plastic Container
293	D25045	Xylometazoline Nasal drop(Paed)	0.1
294		Atropine Ointment	
295		Azithromycin Eye Oint.	1%w/w
296		Benzocaine Ointment	
297	D25011	Chloramphenicol Eye Ointment	1%w/w
298	D25031	Chlorehexidine Oral Gel	1%
299		Clotrimazole +Beclomethazone Skin ointment	
300	D12011	Clotrimazole Vaginal Gel	2%
301	D18030	Cream Clotrimazole	1% w/w
302	D18012	Cream Permethrin Lotion	5%
303	D12020	Cream Terbinafine	
304	LD0060	Diclofenac gel	
305		Gamma Benzene Hexachloride lotation	1%+ Benzocaine 2%
306	D02007	Lidocaine Hcl Gel	2%
307		Lignocaine Spray	
308	D18032	Mupirocin Ointment	2% 5Gm Tube
309		Oint momentasole	
310		Ointment Fluocinolone Acetonide	0.1% ww 30gm tube
311	D38017	Povidone Iodine Mouth Wash	2%
312	D18024	Providone Iodine Oint.	5% w/v , 15Gm Tube
313		Triamcinolone Ointment	0.10%
314		Vit A Cream Oral	
315		Triple layer Face Mask &Head cap	
316		Repsul Flohale	
317		Silk Sutura	Mar-00
318		A.D syringe	0.5ml
319		Alcohol based hand rub as per WHO recommended formula with stand	5 Ltr/ 500 ml
320		Ambu bag Pediatric	250ml
321		Ambu Bag Pediatric	500ml
322		Asthalin Inhaler	
323		B.T SET	
324		Baby weight Scale 5Kg	As per Speciation attached
325		Baciloid Extra	5lt
326	D31018	Black Disinfectant Fluid	4Lit/Jar

327		BP Apparatus (Digital type)	
328		BP Apparatus (Mercury type)	
329		Budesunide +Formotenal Inhaler	
330		Calamine Lotion	
331	S01012	Catgut Chromic 1-0 with needle	Size- 1-0 ,length: 76cm, 1/2 Circle Round bodied ( 45mm needle)
332		Chlorohexidine mouthwash	
333		Cydex Solution *	5ltr
334		Diaper XS Size for newborn	
335		Dimethanol with Gluteral dehyde for fogging	
336		Dispo Syringe	20ml
337		Dispo Syringe	50ml
338		Disposable Hand Care & Face Care	
339	S02010	Disposable Scalp Vein Set-ETD Sterilise Pyrogen free short beveled siliconised needle	size 20, 22, 24, Length 3/4 with CE certification.
340		Drape Sheet Polythene	
341		ECG Papers	
342	S02133	Elastic Adhesive Bandage( fixator)	(IV Cannula Fixator)
343		Elastic crape Bandage	6"
344		Electrical Needle & Syringes destroyer (NST)	
345		Endotracheal Tube	2.5,3,3.5
346		Examination Gloves	Mark-Sizes 6 1/2,7,7 1/2
347	D16048	Factor VII	1.0mg
348		Fluticazone /Rometrol MDI	200/400
349		Fogging Machine	
350		Glutial Dehide	
351		Hand Sanitizer	500ml
352		Hand wash (Soap Base)	200ml
353		HBNC Asha Kit with bag	As per Speciation attached
354		I.V Set Child)	
355		Infant feeding tube	5/6/7/8/10
356		Insulin syringe	40unit
357		intravenous set (Adult)	
358		Lactobacillus sachet	
359		Levosilbutamol MDI	50MCG
360		Levosilbutamol+Budesunide	
361	D02011	Lignocaine Spray Tropical	100mg
362		Lotion Clotrimazole	
363		Machintas Roll	4ftx20mtr
364	D34007	Micropore	
365		Nasogastic Tube ped	5,6
366		Nebulizer	
367	S02019	Operation Gloves	Mark-Sizes 6 1/2,7,7 1/2

368	D21012	ORS	20.5Gm
369	D21029	ORS	4.3Gm
370		Oxygen Mask	
371		Oxygen Tube	
372		Paedl Drip set	
373		Patient Identification Band (Single)	
374		Patient Identification Band Combo (Mother & Child)	
375		Plaster Cutter	
376	LS0025	Plaster of Paris Powder Bag	1 kg-
377		Plaster of paris Roller Bandage 4"	10cmx2.7mtr(Ready made)
378		Plaster of paris Roller Bandage 6"	15cmx2.7 mtr(Ready made)
379		PMO Line 3 way stop cock	
380		Polyamide	
381	D18007	Povidone iodine Solution	500ml/Bottle
382		Probiotic Sachet	
383		Pure Hand (Sanitizer)	500ml
384		Repsul Deoline	
385		Selbutamol Repsule	
386		Soap based handrub with Stand	5 Ltr/ 500 ml
387	LD0069	Sodium Phosphates Enema B.P. 100ml,-	
388		Spiral Needle	(25G)
389		Steam Indicator Tape for checking of Autoclave Sterilization	Tape
390	D19006	Surgical Sprit	500ml
391	S02046	Syringe	5cc
392	S02043	Syringe	10cc
393	S02045	Syringe	2cc
394		Thermometer Clinical	
395		Thermometer Digital	As per Speciation attached
396		Ultrasound Jelly	5kg
397		Vicryl	1,1-0,
398		Weighing Machine (Digital) Adult	120kg
399		Weighing Machine (Digital) Child	5kg

## 2. Mental Health Drugs

Sl No	Name of the item	Speceification	Remarks
1	Tab Amisulpride	100mg/tab	
2	Tab Amitriptyline	25mg	
3	Tab Carbamazipine	300mg/Tab	
4	Tab Carbamazipine	200mg/tab	
5	Tab Chlorpromazine	50mg/Tab	
6	Tab Chlorpromazine	100mg	
7	Tab Clonazepam	1mg/Tab	
8	Tab clonazepam	2mg/Tab	
9	Tab Clonazepam	0.5mg /Tab	
10	Tab Clozapine	100mg/Tab	
11	Tab Diazepam	10mg /Tab	
12	Tab Diazepam	5mg /Tab	
13	Inj Diazepam	5mg/ml	
14	Tab Donepezil	5mg/Tab	
15	Tab Donepezil	10mg/tab	
16	Tab Escitalopram	5mg/Tab	
17	Tab Escitalopram	10mg /Tab	
18	Cap Fluoxetine	20mg/Tab	
19	Inj Fluphenazine	5mg/amp	
20	Tab Fluxoxetine	20mg/Tab	
21	Tab Haloperidol	5mg/Tab	
22	Inj Haloperidol	25mg/amp	
23	Tab Imipramine	75mg/tab	
24	Tab Imipramine	25mg /Tab	
25	Tab Lithium Carbonate	300mg/Tab	
26	Tab lorazepam	2mg/Tab	
27	Inj Lorazepam	1mg/ml	
28	Tab Memantine	5mg/tab	
29	Tab Memantine	10mg/tab	
30	Tab Miltazapine	15mg/Tab	
31	Tab Nitrazepam	10mg/Tab	
32	Tab Olanzapine	10mg/tab	
33	Tab Olanzapine	5mg/Tab	
34	Tab Oxcarbapazapine	150mg/tab	
35	Tab Oxcarbapazapine	300mg/tab	
36	Tab Paroxetine	25mg/tab	
37	Tab Phenytoin Sodium	100mg/Tab	
38	Tab Procydiedin	5mg/Tab	
39	Inj Promethazine	25mg/	
40	Inj Promethazine	50mg	
41	Tab Promethazine	25mg/Tab	
42	Tab Quetiapine	50mg/tab	

43	Tab Quetiapine	100mg/tab	
44	Tab Reperidone	0.5mg	
45	Tab Risperidone	2mg/Tab	
46	Inj Risperidone		
47	Syrup Sodium Valporate		
48	Tab Sodium Valporate	500mg	
49	Inj Thiamine	10mg/	
50	Tab Trihexyphenidyl	2mg/Tab	
51	Tab Uthium Carbonate	300mg	
52	Tab Carbamazepine	200mg	
53	Tab Diphenhydramine	100 mg	
54	Inj Fluphenazine	25mg	
55	Cap Fluxoteine	20mg/Cap	
56	Tab phenobarbitone	30 mg	
57	Tab phenobarbitone	60 mg	
58	Piracitam Syrup		
59	Tab Zolpidem	10mg	

### 3. LIST OF NCD Drugs.

Sl No.	Drugs code	Drugs Name	Remarks
	<b>Anti-Hypertensive Medicines</b>		
1	Non-EDL	Tab. Amlodipine 2.5 mg	
2	D17015	Tab. Amlodipine 5.0 mg	
3	NA	Tab. Amlodipine 10.0 mg	
4	D17006	Tab. Atenolol 50 mg	
5	NA	Tab. Atenolol 100 mg	
6	D17036	Tab. Enalapril 2.5 mg	
7	D17049	Tab. Enalapril 5.0 mg	
8	D17032	Tab. Telmesertan 20mg	
9		Tab. Telmesertan 40mg	
<b>Cardiovascular Medicines (Medicines used in Angina)</b>			
10	D17001	Tab. IsorbideDinitrate 5.0 mg	
11	D17001	Tab. IsorbideDinitrate 10.0 mg	

*B*

12	D17048	Tab. Clopidogrel 75 mg	
13	D2001	Tab Furosemide 40 mg	
14	NA	Furosemide Oral Liquid 10mg / ml	
15	D 17026	Tab Hydrochlorothiazide 12.5 mg	
16	D20005	Tab Spironolactone 25 mg	
17		Tab Spironolactone 50 mg	
<b>Anti-Diabetic Drugs</b>			
18	D22019	Tab. Glimiperide 1 mg	
19	D22014	Tab. Glimiperide 2 mg	
20	D 22002	Inj. Insulin (Soluble) 40 IU / ml	
21	NA	Inj Insulin (Intermediate Acting 40 IU / ml (NPH)	
22	D22012	Inj Insulin Premix (30:70) (Regular NPH) 40 IU/ ml	
23	D 22003	Tab Metformin 500 mg	
24	NA	Tab Metformin 1000 mg	
<b>Anti-Convulsants / Antiepileptics</b>			
25	D07003	Tab Carbazepine 200 mg	
26	D07003	Tab Carbazepine 300 mg	
27	NA	Tab Diazepam Oral liquid 2mg / 5 ml	
28	D07001	Tab Phenobarbitone 30 mg	
29	NA	Tab Phenobarbitone 60 mg	
30	D07011	Phenobarbitone Oral liquid 20 mg / 5 ml	
31	D07018	Tab Phenytoin 50 mg	
32	D07004	Tab Phenytoin 100 mg	
33	NA	Tab Phenytoin 300 mg	
34	NA	Tab Phenytoin ER Tabs 300 mg	
35	NA	Inj. Phenytoin 25 mg / ml	
36	NA	Inj. Phenytoin 50 mg / ml	
37	D07005	Tab Sodium Valproate 200 mg	
38	D07022	Tab Sodium Valproate 500 mg	
<b>COPD</b>			
39	D28011	Tab Salbutamol 2 mg	
40	D28004	Tab Salbutamol 4 mg	
41	D28011	Salbutamol Oral liquid 2 mg / 5 ml	
42	D28018	Salbutamol for Nebuliser 5 mg / ml	
43	D28013	Salbutamol Inhalation (MDI / DPI) 100 mcg / dose	
<b>Others</b>			
44		Nicotex tablet under NTCP 2mg	
45		Nicotex tablet Under NTCP 4mg	
46		Glucometer for diabetes	
47		Blood Lancet	
48		Glucostrips	
49		Clotrimazole Mouth Paint (Plastic Container as per IP)	
50		Soln. Formaldehyde (Plastic Container as per IP)	
51		Benzylamine HCl Mouth wash	

52	Chlorhexidine Gluconate solution (mouth wash)
53	Analgesis, Anti Inflammatory & Anti bacterial
54	Clotrimazole & Lignocaine Ear Drop
55	Fluticasone Nasal Spray
56	Lidocaine + Ofloxacin Ear Drop
57	Mometasone nasal spray
58	Oxymetazoline Nasal Drop
59	Saline Nasal Drop
60	Wax Softner
61	Xylometazoline Nasal Drop
62	Eat safety Magic Box

#### 4. LIST OF X-RAY & NIDAN LABORATORY CHEMICALS, ANTIGEN, PIPETTE, VIALS & CONSUMABLES

Sl no	Name of the items
1	Glucose kit
2	Serum urea kit
3	Serum creatinine kit
4	Serum cholesterol kit
5	Serum Bilirubin kit
6	ASO titer Antigen
7	R.A Factor Antigen
8	CRP Antigen(tulip)
9	Toxo plasma strip(tulip)*
10	VDRL Strip(SD)
11	Disposable ESR Pipet(with vacuum plug)
12	Anti A Group tulip
13	Anti B Group tulip
14	Anti D Group(igg&igm)tulip
15	Micro tips 100 micron
16	Micro tips 1000 micron
17	Sodium Citrate For E.S.R Test 3.2% & 3.8% 500 bottle
18	Sodium Fluoride Vial (Double Oxalated)
19	E.D.T.A Vial
20	Plain vial
21	Urine Container
22	Dist Water
23	Malaria (ICT) Kit
24	Lancet
25	Glucostick For sugar(Code Free) (SD)
26	Urostick
27	Hemoglobin Tube& pipette
28	Widal Antigen

29	Lishmans Stain
30	Benedicts Salutation
31	Micro Slides
32	Heamometer Set
33	WBC Diluting Fluid
34	HB SAG, HCV kit
35	Benzidine Powder
36	Heparinised Capillary tube for Micro ESR
37	Sulphor powder
38	Barium Chloride
39	Glass Test tube
40	Glacial acetic acid
41	Cover Slip
42	N/10 HCL -500ml.bott.
43	X-Ray Fixer
44	X-Ray Developer
45	Sodium Hypochlorite (10 %) 5 ltr./jar
46	Disposable Apron
47	Plastic/ Rubber Apron
48	Disposable Cloth for wrapping the newborn babies
49	Plastic bag for drying BWB polythene
50	U.S.G. Jelly 5ltr./jar
51	Stethoscope 3 Different Brands.
52	Oxygen Cylinder Stand (for 1.5 Liter cylinder)
53	Oxygen Flow meter with Humidifier
54	Slide Wrench

55	Oxygen Key
56	Dimethanol with Gluteral Dehyde 5 Ltr Jar
57	Alcohol based hand rub as per WHO recommendation
58	Oxygen Cylinder 1.5 cumm
59	Digital Hemoglobinometer
60	Digital Hemoglobinometer(Stick)
61	Occult Blood Test kit
62	HIV Test Kit
63	Uric Acid Kit
64	ECG Paper roll 208,2mmx 20mtr
65	Sodium Citrate Vial For E.S.R Test 3.2% & 3.8%
66	Urine Pregency test kit
67	Colori meter test tube
68	Test tube Washing Brush
69	Tissue Paper
70	Plastic Dropper 2ml
71	Plastic Dropping Bottle 100ml
72	Silanized Glass Pipette
73	Slide tray
74	Test tube stand
75	3.8% sodium Citrate 500ml bottle
76	Staining Glass Jar
77	Immersion Oil
78	White Pencil
79	Glass Marking Pencil
80	Platelet Dulain Fluid
81	Torniquit (Flemigo)
82	Test tube Glass
83	Test Tube Holder
84	Liquid Parafin Light
85	Glass Beaker 500ml
86	Glass Beaker 250ml
87	Trucks Fluid (WBC)
88	Hems Fluid(RBC)
89	Capillary Tube for BTCT
90	Thermocool Box (12"x12"x12")
91	Ice Gel Pack
92	K3 EDTA vacuum Vial
93	Sodium Citrate vial 3.8%non-vacuum vial (Black Cap)
94	Pregency test Kit
95	Blood clot Activator with gel non vacuum
96	Non vacuum plain vial

### 5. LIST OF BEDDING, CLOTHING, AND LINEN ITEMS

Sl no	Name of the items	Specification	Remark
1	Bed sheet (white coloured, bleached, both side stitched)	Length=230 Width =150cm Construction particulars, Performance requirement dimension and mass should confirm to IS 175:1989 (Reaffirmed 2001, Variety No.4)	
2	Turkish Towel (Large white colour)	150cmx70cm.	Both side stiched
3	Turkish Towel (Small, White colour.)	60cmx80cm.	Both side stiched
4	Green Cloth cotton	89cm/Mtr	
5	O.T. Towel (Green colour)	100cmx89cm	Cotton, both side stiched
6	Pillow Cover		
7	Pillow		
8	Blanket (Adult)		
9	Mosquito Net		
10	Patient Gown		
11	Surgen Gown		
12	Mackintos Sheet	(20mtr Roll)	
13	Mattress Cover		

## 6. List of DPH Laboratory Items:

DPH, KENDRAPARAI Lab Logistic & Consumables			
Glass wares			
Sl. No.	Name of the Apparatus	Quantity	Specification
1	Conical Flask	100ml	Glass
		250ml	
		500ml	
		1000ml	
2	Measuring Cylinder	10ml	Glass
		100ml	
		250ml	
		500ml	
		1000ml	
3	Petriplates	100mm	Glass
		100mm	Autoclavable fiber
		100mm	Disposable
4	Test Tube( Rimless)	12x75	Glass
		10ml	
		20ml	
		38ml	
		30ml	
		100ml	
5	Beaker	50ml	Tarson Autoclavable Plastic
		100	
		250	
		500ml	
		1ltr	
6	Borosil Microscope Slide	Transparent	Glass
7	Autoclavable Glass Bottles with Lid	500ml	Autoclavable
8	Coverslip	Transparent	Pkt
9	Durham's Tube	100/pkt	Glass
10	Concavity slide	Glass	for Motility testing

Metalwares			
Sl. No.	Name of the Apparatus	Quantity	Specification
1	Inoculating Loop	1mm Diameter	Nichrome(Himedia)
		1mm Diameter	
2	Forceps	Medium	Stainless Steel
3	Scissors	Medium	Stainless Steel
4	Staining Rods		Steel
5	Staining Rack		Steel

6	Spatula	Medium
---	---------	--------

Test kits			
Sl. No.	Name of the Test kit	Quantity	Specification
1	Dengue IgM ELISA Kit		Meril
2	Dengue NS1 ELISA Kit		Meril
3	H2S( Coliform)		Himedia
4	Orthotoludine kit		Himedia
5	JE IgM ELISA Kit		
6	HBA1C Kit		5D
7	Leptospirosis IgM ELISA Kit		ELISA kit
8	Scrub Typhus IgM ELISA Kit		ELISA kit
9	Chikungunya IgM ELISA Kit		ELISA kit
10	Measles IgM ELISA Kit		ELISA kit
11	Quantitative viral load kit for HBV		TRUE NAT KIT
12	Quantitative viral load kit for HCV		TRUE NAT KIT
13	Hepatitis A Ig M		ELISA kit
14	Hepatitis E Ig M		ELISA kit

Other requirements			
Sl. No	Name of the apparatus	Quantity	Specification
1	PH paper		small Range wide range
2	Sample Collection Vials	Plastic	Capped(plastic)
3	Sample Container	Plastic	Sterile Uricol
4	Cotton Swabs	Sterile	(In Packets)
5	Filter Paper	Wattman's	
6	Aluminum foil		
7	Thermo cool box		10x10x10 Inchs
8	Ice gel pouch		
9	Spreader	In packet	
10	White Pencil Marker		
11	Waste disposal Bags( Black, Blue, red, Yellow)	50/packet	
12	Zip lock bag plastic		
13	Discarding Tray(stain less steel)	Hard Plastic	Autoclavable
14	Sterilizing Indicator(Sterib thermophilus strip)	Strips in Packet	
15	Brown paper		
16	Cello tape Big size		
17	Rubber bands	In KG	
18	Tissue paper	In Roll	
19	Cotton	Absorbant	In Packet

		Nonabsorbent	In Packet
20	Multichannel Micropipettes(Adjustable)		20-300 µl
		1000/pkt	10 µl
21	Sterile Micro Tips(for ELISA)	1000/pkt	200 µl
		1000/pkt	1000µl
22	Aprone	White	Full sleeve
23	Nitrile Gloves( Blue colour sterile powder free)	50/Pkt	5.0"
		50/Pkt	7.0"
24	Masks disposable	100/Pkt	disposable
25	Antibiotic Zone Diameter Scale	Plastic	
26	Test Tube Washing Brush		Big medium small
27	K3 EDTA vials		
28	Plain vials		
29	Gel clot vials		
30	Sterile swab stick	.....	Himedia(PW1136)
31	Sterile swab stick (Screw cap)	.....	Himedia(PW003)
32	Tourniquet		
33	10 ml screw cap vials		Tarson
34	Autoclavable Sterile Plastic Vials	5ml	Plastic
35	Autoclavable Sterile Plastic Vials	10ml	Plastic
36	Urine culture Bottle	50 ml	
37	Water testing bottle(Plastic Autoclavable)	250ml	Tarson
38	Bunsen Burner with Pad & Stand		
39	Disposable Syringe	2ml 5ml 10ml	

#### Liquid Reagents

Sl. No.	Name of the Liquid reagent	Quantity	Specification
1	Cedar wood Oil/Liquid Immersion Oil	in Bottle	100ml
2	Absolute Alcohol (99.99%)		500ml
3	Isopropyl Alcohol (70%)		
4	Spirit(Surgical)		500ml
5	Lysol	in Lt.	
6	3% Hydrogen peroxide(H <sub>2</sub> O <sub>2</sub> )		500ml
7	Kovac's Indole reagent		
8	Barrit's reagent A		
9	Barrit's reagent B		
10	KOH		500gm
11	Sodium Hydroxide(NaOH) crystals		100gm
12	4% Sodium Hypochloride solution		lit Can

13	Formal Dehyde		500ml
14	Concentrated Sulphuric Acid	99.99%	250ml
15	Acetone		500ml
16	Buffer Capsules(pH 4& 9)		1pkt
17	Deionised/Distilled water	In Ltr	
18	Dettol Hand wash		
19	Deoderising Pearl		
20	Urea Crystals for Urease Agar		250gm
21	Diluent for CBC 3 part		Medonic
22	Lyse for CBC 3 part		Medonic
23	Liquid Paraffin		200ml
24	Sodium Chloride		
25	Sodium Hydroxide		
26	HCL		
27	Sanisol Disinfectant	In ltr	
28	10% Feric Chloride		
29	Phenol Indicator		

Media Required			
Sl. No.	Name of the Culture Media	Bottle containing	Specification
1	Nutrient Broth	500mg	Himedia
2	Nutrient agar	500mg	Himedia
3	Agar Agar	100gm	Himedia
4	Peptone Water	500mg	Himedia
5	Mac-conkey broth	500mg(Single strength)	Himedia
		500gm(Double strength)	Himedia
6	Mac-conkey Agar	500mg	Himedia
7	Selenite F Broth	200gm	Himedia
8	XLD Agar	200gm	Himedia
9	DCA agar	200gm	Himedia
10	TCBS agar	500mg	Himedia
11	Mullar Hinton Agar	500gm	Himedia
12	Triple Sugar Iron Agar	500gm	Himedia
13	Simmon citrate Agar	500gm	Himedia
14	Carry Blair medium in polypropylene tube		Himedia
15	Alkaline peptone water	500mg	Himedia
16	Lysine Iron Agar	500gm	Himedia
17	Mannitol Motility Agar	100gm	Himedia
18	Salmonella Shigella Agar	500gm	Himedia
19	Brilliant Green Bile Broth	500gm	Himedia
20	MR-VP broth	500gm	Himedia
21	Urea Agar Base(Christensen)	500gm	Himedia

22	Loeffler's serum Medium	100gm	Himedia
23	Streptotrophomonas selective Agar Base	500gm	Himedia
24	Methyl Red Indicator		Himedia
25	Mannitol Salt Agar	100gm	Himedia
26	Robert Cook meat broth	500gm	Himedia
27	Robert Cook meat broth(Ready made)		
28	Sheep Blood Agar 90 mm-plate (Ready made)		Himedia
29	Brain Heart Infusion Broth bottle ( Ready made) for pediatric.	20 ml	Himedia
30	Brain Heart Infusion Broth bottle (Ready made) for Adult.	50 ml	Himedia
31	CLED Agar	500gm	Himedia
32	Carry blair media base	500gm	Himedia

#### Stain & Antibiotics

Sl. No.	Name of the Stain kit	Specification
1	Gram staining kit	Himedia
2	AFB Stain Kit	
3	Leishman's Stain kit	Himedia
4	Oxidase disc Bottle	Himedia
5	Albert Stain Reagent(1&2) Kit	Himedia

#### ANTIBIOTICS DISC REQUIRED

1	Amikacin 30 mcg	100x1 Himedia
2	Ampicillin Ab 10 mcg	100x1 Himedia
3	Azithromycin 15 mcg	100x1 Himedia
4	Amoxicillin clavulanic acid 18 mcg	100x1 Himedia
5	Aztreonam 22 mcg	100x1 Himedia
6	Cefuroxime 18 mcg	100x1 Himedia
7	Ertapenem 22 mcg	100x1 Himedia
8	Imipenem 23 mcg	100x1 Himedia
9	Mertapenem 23 mcg	100x1 Himedia
10	Cefoperazone 75 mcg	100x1 Himedia
11	Cefixime 5 mcg	100x1 Himedia
12	Cefodoxime 10 mcg	100x1 Himedia
13	Cefoperazone 75 mcg	100x1 Himedia
14	Cefotaxim 5 mcg	100x1 Himedia
15	Ceftriaxone 30 mcg	100x1 Himedia
16	Ceftazidime 21 mcg	100x1 Himedia
17	Chloramphenicol 30 mcg	100x1 Himedia
18	Ciprofloxacin 5 mcg	100x1 Himedia
19	Clindamycin 2 mcg	100x1 Himedia
20	Co-Trimoxazole 25 mcg	100x1 Himedia
21	Doxycycline 30 mcg	100x1 Himedia

22	Erythromycin 15 mcg	100x1 Himedia
23	Gentamycin 10 mcg	100x1 Himedia
24	Levofloxacin 5 mcg	100x1 Himedia
25	Linzolid 30 mcg	100x1 Himedia
26	Minocycline 16 mcg	100x1 Himedia
27	Nitrofurantoin 100 mcg	100x1 Himedia
28	Nitrofurantoin 100 mcg	100x1 Himedia
29	Norfloxacin 10 mcg	100x1 Himedia
30	Ofloxacin 5 mcg	100x1 Himedia
31	Oxacillin 1 mcg	100x1 Himedia
32	Piperacillin 100 mcg	100x1 Himedia
33	Tobramycin 10 mcg	100x1 Himedia
34	Tetracycline 15 mcg	100x1 Himedia
35	Vancomycin 30 mcg	100x1 Himedia
36	Cefuroxime	100x1 Himedia

Sl. No.	Item Description	Pack Size	Make	Cat No.
1	Pipettor Stand 5 places	1	Tarsons	30080
2	Cryo Cube Box, 100 places, 4box/pack	1.8ml	Tarsons	202070
3	Cryo Cube Box, 100 places, 4box/pack	1.8ml	Tarsons	212070
4	Micro Centrifuge Tube, 500/pkt	2 ml	Tarsons	500020
5	Purple Nitrile Gloves L size, 50/pkt	Large	Tarsons	97611
6	Purple Nitrile Gloves M size, 50/pkt	Medium	Tarsons	97612
7	Racked Filter Tips, Sterile, 960 (10 Racks)	1000 ul	Tarsons	528106
8	Racked Filter Tips, Sterile, 960 (10 Racks)	200ul	Tarsons	528104
9	Racked Filter Tips, Sterile, 960 (10 Racks)	20ul	Tarsons	528101
10	Racked Filter Tips, Sterile, 960 (10 Racks)	10ul	Tarsons	528100
11	Ethanol Absolute	500ml	Adventol	
12	Reversible Rack with Cover, 4/pkt	0.5 or 1.5ml	Tarsons	241010
13	CRYOCHILL- TM -20°C Mini Cooler, 12 Places	1.5ml	Tarsons	525010
14	Polygrid Micro Centrifuge Rack		Tarsons	241014
15	RNA Erase	250ml	Himedia	ML162
16	Parafilm M, 2"x250'		Tarsons	380010
17	Parafilm M, 4"x125'		Tarsons	380020
18	Disposable Apron			
19	Optical Plate Seal, 100/pkt		Tarsons	157300

20	Semi Skirted 96 Wells x 0.2 ml Plate, 25/pkt		Tarsons	612020
21	PCR Rack with Cover, 6/pkt		Tarsons	241000
22	Indicator Tape for Steam Autoclave, 1"x500"	1 roll	Tarsons	670050
23	Autoclavable Biohazard Bags, 12"x24", 100/pkt		Tarsons	550012
24	Autoclavable Biohazard Bags, 24"x30", 100/pkt		Tarsons	550015
25	Cryo Tags, 38x19, 1000/pkt		Tarsons	523090
26	Tough Tags, 32.5x12.7, 1000/Pkt		Tarsons	500060
27	Multipurpose Labelling Tape, 075"x500"		Tarsons	680000
28	SPINWIN -TM Centrifuge Tube Conical Bottom, 300/pkt		Tarsons	546042
29	Wash Bottle-500ml Capacity		Tarsons	560030
30	Measuring Cylinder 6/Pkt, 250 ml		Tarsons	345050
31	Measuring Cylinder 6/Pkt, 500 ml		Tarsons	345070
32	Test Tube Stand, 12place, 32mm dia, 4/pkt			202050
33	Test Tube Stand, 20place, 20mm dia, 4/pkt		Tarsons	202030
34	Water nuclease-free, 30 mL		Genetik	80582
35	Cryogenic Marker		Tarsons	930011
36	Cryochill TM Vial Self standing sterile, 1000/pkt	1.8ml	Tarsons	523182
37	UNIVERSAL MICRO TIP BOX, 10/pkt	0.2-10 ul	Tarsons	524052
38	UNIVERSAL MICRO TIP BOX 10/pkt	2 - 200 ul	Tarsons	524053
39	UNIVERSAL MICRO TIP BOX 10/pkt	200-1000 ul	Tarsons	524059
40	Aluminium foil			
41	Maxlamp 0.1 ml tube strips with cap		Tarsons	611010
42	Beaker, PP Autoclavable, 6/pkt	500ml	Tarsons	421050
43	Beaker, PP Autoclavable, 12/pkt	250ml	Tarsons	421040
44	Biohazardous Waste Container	5 lit	Tarsons	583256
45	Biohazardous Waste Container	20 lit	Tarsons	583254
46	Biohazardous Waste Container	1.5 lit	Tarsons	882209
47	Cryo Box Rack/PCR Plate holder, 14x14x60cm, 2/pkt		Tarsons	524034
48	Tissue Paper			
49	Rack for Micro Tube, 48 places, 6/pkt	1.5ml	Tarsons	241030
50	Spillyter Lab Soakers, 250pads/pkt		Tarsons	590010
51	Kim wipes		Tarsons	
52	Parafilm Dispenser	1pc	Tarsons	380000
53	AlkaSept	1ltr Pack	Himedia	C0152

54	Accupipette-T2-0.5 ul – 2ul	Tarsons	30000
55	Accupipette-T20-2 ul – 20ul	Tarsons	30010
56	Accupipette-T10-10 ul – 100ul	Tarsons	30030
57	Accupipette-T200-20 ul – 200ul	Tarsons	30040
58	Accupipette-T1000-200 ul – 1000ul	Tarsons	30050
59	MAXAMP PCR Tube 0.2ml	Tarsons	510051
60	Cryobox (1.8ml Capacity, 4/Pkt)	Tarsons	524020
61	96 PCR Tubes Thermo Conductive Rack	Tarsons	525320
62	Utility Tray	Tarsons	242010
63	Utility Tray	Tarsons	240070
64	Filter Pipette Tips 30ul Capacity	Tarsons	527101
65	Filter Pipette Tips 200ul Capacity	Tarsons	527104
66	Filter Pipette Tips 1000ul Capacity	Tarsons	527106
67	Micro Tips 1000ul Capacity	Tarsons	521020-8
68	Sample Bag 5"x8"	Tarsons	553020
69	Multipurpose Labelling tape	Tarsons	580000
70	Spinwin Microcentrifuge Strip MP 100	Tarsons	

#### 7. LIST OF LABORATORY MATERIAL REQUIRED FOR RNTCP USE

Sl NO	Name of the Item	Name of the manufacturer	Pack Size	Specification	Remark
1	Distil Water		5ltr. Jar	RNTCP Standards	
2	Basic Fuchion	Himedia, Merck, Rankem,	25Gm/Bottle	Chemical Name: Paraosaniline Hydrochloride, Chemical Formula: $C_{20}H_{20}ClN_3$ , Mol. Wt:- 337.86, Dye Content: Approx. 85%-88%, Colour: Metallic Green.	
3	Phenol	Himedia, Merck, Rankem,	500ml/Bottle	Chemical name: Phenol, Chemical structure: $C_6H_5OH$ , Molecular Wt: 94.11, Melting point: $40^\circ C + 2$ , purity: 99.5%	
4	Methylene Blue	Himedia, Merck, Rankem,	(25Gm/Bot)	Chemicals name: Methyl thionine chloride, Chemical structure: $C_{16}H_{18}ClN_3S$ , Molecular Wt: 319.5, Dye content: Approx 82% (dye content must be mentioned)	

5	Conc. Sulphuric Acid	Himedia, Merck, Rankem,	500ml/Bottle	Chemicals name :Sulphuric Acid, Formula Wt: 98.08, Specific Gravity 1.84 Minimum Assay 98%
6	Synthetic Oil/Immersion oil for microscopy	Himedia, Merck, Rankem,	30 ml/Bottle	RNTCP Standards
7	Diamond Pencil		Lab Use for Glass	6" holder with artificial diamond (hard stone) embedded at one end with screw cap to mark on microscope glass slides.
8	Glass rod		Big Size(min 3')	Glass rod
9	Waste Bucket		per piece	foot operating plastic bucket with cover.
10	Filter Paper		100 pcs/pack	Filter paper No. 1, 12.5 cms diameter for filtering carbol fuchsin using a small funnel of 10cm maximum diameter, Smooth on one side and rough on the other side, water permeable.
11	Tissue paper		Rolling Type	RNTCP Standards
12	Lens paper		Book Type	Soft microscope lens cleaning tissue, 4" x 6" booklet, each booklet containing 100 sheets
13	Sticker for sputum container		1000 Pack	RNTCP Standards
14	Glass slide	Himedia, Blue Star	50/Pack	Glass microscope slides, 75mm x 25mm x 1.35mm, plain, one pack of 50 slides.
15	Phenyl 40%	Bengal chemicals	5lit	Household disinfectant, containing phenolic compounds such as monochlorophenol, coal tar acid, oils & Emulsifiers etc. The approx content of phenolic compounds should be at least 40%.
16	Plastic dropping bottle		100ml	250ml capacity built-in stem all polypropylene; leak proof screw cap made of polypropylene.
17	Methylated spirit		500 ml	Chemical Name: Ethanol denatured + 5% Isopropyl alcohol + 5% Methanol, Molecular structure: C <sub>2</sub> H <sub>5</sub> OH, Molecular wt:

*B*

				46.07, Purity: 90%
18	Microscopic bulb 12 volt		Philips	
19	Slide Box 100nos rack capacity			Slide Box 100nos rack capacity
20	Slide box 50nos rack capacity			Slide box 50nos rack capacity
21	Falcon tube	Tarson,		
22	Parafilm	Tarson,		Parafilm, 100 mm width, 38" length, for use between 40°C to +50°C
23	N95 Mask			ISI marked respirators, universal size, less than 0.22 µm, individually packaged (preferably in boxes of 10 pieces each)
24	Absorbable Cotton			Absorbent, each roll of 0.5 kg.
25	Disposable Gloves			Acid & alkali proof, Rubber, Anti-static, to cover at least 1/4th of forearm.
26	ENAMEL TRAY			Size: 15" length, 10" Width, 3" Depth
27	Vaccine Carrier			
28	Sputum Container (5ml)			Cups made of Special Medical Grade Polypropylene, thin plastic, translucent, Diameter - 4cm, Capacity - 30 ml, Screwable Cap should also be made of Special Medical Grade Polypropylene and should be air tight. Leak proof.
29	Glass Funnel		Big Size(200ml)	
30	Glass Funnel		Big Size(100ml)	
31	Falcon Tube Stand		20tubes/stand	
32	Gel Pack			
33	Brown Tape			
34	Thermo cool Box			
35	Hygrometer Digital			Hygrometer Digital
36	Weight Machine (For Patients)		Min 200kg capacity	Min 200kg capacity
37	Auramine O Powder,	Himedia,Merck,Rankem,	500g	Chemical formula- C <sub>17</sub> H <sub>22</sub> ClN <sub>3</sub> , Mol. Wt: 303.84
38	Pottassium permanganet Powder	Himedia,Merck,Rankem,	500g	Chemical formula- KMnO <sub>4</sub> , Mol. Wt:158
39	Ethanol(99%),		500ml/Bottle	Ethanol

40	HCL Concentrated,	Himedia,Merck,Rankem,	500ml/Bottle	Chemical name- Concentrated Hydrochloric Acid, Chemical formula- HCl, Mol Wt:36.46, Specific Gravity 1.18
41	Oxalic Acid,		500g	Chemical formula, COOHCOOH.2H2O; Mol Wt:126.07 ; purity 99.8%
42	Measuring Cylinder 100ml	Borosil		Graduated 100ml cylinder
43	Measuring Cylinder 1000ml (Borosil)	Borosil		Graduated 1000ml cylinder
44	Flat Bottom Flask, 500ml (Borosil)	Borosil		Graduated 500ml cylinder
45	Flat Bottom Flask, 2ltr (Borosil)	Borosil		Autoclavable and dry heat resistant, flat bottom, conical flask.
46	Screw Capped Amber Colour bottle, 2ltr	Borosil		Screw Capped Amber Colour bottle, 2ltr. Capacity.
47	Screw Capped Amber Colour bottle, 500 ml	Borosil		Screw Capped Amber Colour bottle, 500ml Capacity.
48	Aluminum Foil for covering auramine regent			Aluminum Foil for covering auramine regent
49	Apron XL			
50	Absorbent Paper			
51	Spray Head Bottle			1 litre bottle with spray head (preferably with one spare spray head)
52	Thermo meter (Alcohol based)			
53	Falcon Tube Stand			
54	Surgical tray			
55	Weight Machine (For Labs)		Min 50g capacity	Min 50g capacity
56	Kim Wipes			
57	Sodium Hypochlorite (4%)			Chemical formula: NaClO, Concentration not less than 4%, Packaging -5 Liter/pack
58	Induction Heater		1000 Watt	

The lab reagents should be procured in good quality from reputed firms like HIMEDIA/ RANBAXY/MERK maintain quality of RNTCP with the potency, mfg Date & Exp Date clearly mentioned on the bottle.

Sample Verification of the Item(s):

Sample verification of the item is a part of the technical evaluation. Before opening of the Price Bid, the sample of the item(s) submitted by the technically qualified bidders (based on document submitted) shall be verified by the Tender committee/Technical teams in order to verify the quality standard as asked in the technical specification. The Tender Committee shall examine the sample and also verify the dimensional parameters mentioned in the technical specification.

Failure to submit the samples before the stipulated date of sample submission or if the samples submitted will be found to be not as per technical specification will lead to automatic rejection of the bid and such bidders shall not be considered for opening of their price bids.

### **8. List of Items required for DEIC of kendrapara district**

Sl No	DEIC Unit	Name of the Equipment	Remarks
1	Vision equipment	Log Mar Charts	
2		Snellen's Chart	
3		Pediatric Trial Lens Set	
4		Pencil torch with yellow light	
5		children's Trial Frame	
6		Vision testing drum (Distance)	
7		Illuminated Nar Vision Test Drum	
8		Infant and toddler VEP or Visual Evoked Potential Testing equipment	
9		Lea Grating Paddles	
10		Lea Near Vision Single symbol playing cards	
11		Lea Near vision cards with symbols with card	
12		LEA symbol 13- Line Translucent Distance Chart (10 feet/3 meters)	
13		LEA Symbols Low contrast 10m Flip Chart	
14		Lea 3-D Puzzle	
15		Hiding Heidi Low contrast Face Test, Serial no- 253500: Double sided	
16		Heidi Expressions Test Game serial noo, 254505	
17		Hiding Heidi Low contrast Face Test, Serial no:- 253500: Double Sided	
18		Spectacle Occluder item # 52327	
19		Occluder Glasses item #52851	
20		AAPCS Vision Screening Kit item no.# 120	
21		LEA SYMBOLS Chart for Vision Rehabilitation Item no-258000	
22		Leo Learns by doing DVD for interaction with a vision impaired child (79000)	
23		COVID Pediatric Kit - 400444 SKU number	
24		Lea Fixation Stick	
25		Canon Power shot TX1 digital camera	

26		Random Dot Butterfly Stereopsis test with intermediate sized polarized glasses	
27		Receptive Expressive Emergent Language Test - Third Edition (REEL-3)	
28		LPT Linguistic Profile test	
29		Developmental assessment for Indian Infants (DASSI)	
30		Vineland Social Maturity Scale	
31		Vineland Adaptive Behaviour Scales	
32		Developmental Screening Test (DST) by Bharat Raj	
33		Denver Developmental Screening Test II (DDST-II)	

34		Stanford Binet (Indian adaptation Kulshresta)	
35	Equipments for Psychological Assessments	bayley - III Screening Test complete kit Includes: Manual , Stim Book, Picture Book, Record Forms 25 Packs	
36		Piagets Sensori-motor intelligence scale 0-2 Years "Piagetian cognitive Tasks"	
37		Dyslexia Early Screening Test 4-6 Years (DEST)	
38		Dyslexia Early Screening Test Junior (6-11 Years)	
39		Childhood Behavioral Checklist CBCL	
40		Cerebral Palsy and Neuro-Motor impairment ; INCLIN (INDT-NMI)	
41		INCLIN Diagnostic Tool for Epilepsy (INDT-EPI)	
42		ADHDT2	
43		SANGUIN FORM BOARD	
44		Pinspot and Mirror Ball	
45		Mirror Ball Motor	
46		LED Mirror Ball	
47		Fire Ball mounted on the roof	
48		Sound Activated Light	
49		LED Bubble Tube	
50		OPTIC Fibers	
51	Sensory	Bubble Tube	
52	Integration	Roting Drum	
53	Equipments	Chime Frame and Beater	
54		Mirror Chime Bout	
55		Swings:	
56		Boister Swing	
57		Platform Swing	
58		Tyre tube Swing	
59		Rope ladder Swing	
60		Rhythmic Rocker	
61		HIPRO	
62		OPD material	
63		Thudicum's speculum set	

64	Killian speculum set	
65	Laryngeal Mirrors set	
66	Posterior rhinoscopic mirrors	
67	Aural speculum all sizes black finish	
68	Nasal packing forceps	
69	Aural Suction Tips (Micro suction adapter)	
70	Nasal suction Tips	
71	Tuning Fork 1 set-1 tuning fork of 4 different freq 128 Hz, 256 Hz, 512Hz and 1024 Hz	
72	Hartman's Forceps	
73	Eustachian tube catheter	
74	Punch biopsy forceps	
75	Laryngeal biopsy forceps	
76	BP Handles of various sizes (Available in district Store)	
77	Tongue depressor	
78	Therapy ball :65 cm	
79	45 cm and a pump for inflation	
80	Therapy mats- 6ft x 3ft (Length 6ft and Breadth 3ft. Thickness 4cm	
81	Boister	
82	2ft Long, diameter 8 inch	
83	2ft long , diameter 10inch	
84	Small Roll-13 inch long, diameter-3inch	
85	<b>Prone wedges</b> Foam filled wedges covered with Nylon fitted with Rexene straps to position the child	
86	Bug Height -14 inch Length - 31 inch, Braedth - 17 inchs	
87	Small Height -10 inch Length -26 inch.breadth-17 inches	
88	Balance Board Rexene covered sushioned Platform, size 45cm x 60cmx15cm high	
89	Kaya Walker (height-48-64 cm distance between hand grips 34 cm frame width 58-60cm, frame length 69-83 cm, user height 107- 137 cm, maximum user weight 39kg, frame weight 3.85 kg)	
90	Trampoline (small)	
91	Boister Swing With nylon rope or straps with hooks to fir in the swing frame , Size 25 cm diameter X 90 cm Long	
92	Wooden Benches with cushion and Rexene cove- a) small (3ft long, height 8 inches, breath 6 inches)	
93		
94	b) Big (3ft Long, height 12 inches, breath 8 inches)	
95	Splints (Ankie Foot orthosis)	
96	Special Chair With Cut - out	

		(tray) Tailor made according to need of the child)	
97		Cut - out floor table (2ftx2ft)	
98		Floor seat (Pelvic strap)	
99		locking mirrors	
100		Toys (for play and stimulation)	
101		Small rattles	
102		Squeaky toys	
103		Puja bell (clapper bell)	
104		Soft toys	
105		Brush for tactile stimulation	
106		Therapeutic Gluten free, non - toxic, red, yellow and blue colors	
107		Peg board laminated square board having 10 holes to hold smoothly finished solid plastic pegs in five different bright colours	
108		Ball pool, The dense foam padded mini Ball Pool is soft, safe and perfect for small children. It provides an excellent sensory stimulating activity. The round pool is 120 cm in diameter x 50cm high, & has 10cm thick padded sides. The pool contains 500 multi color balls of 7cm or 8cm diameter. Pool side and bottom is covered with durable Rexene that easily wipes clean.	
109		Balls of different size	
110		Galters Aluminium/bamboo sticks of 8", 10", 12", 14" long inserted in the pockets of thick canvas, 3 Velcro straps to be wound around Total 8 nos (1 pair of each size mentioned)	
111		Thick handle spoon stainless steel, spoon, padded handle 3	
112		Thick handle bent spoon Stainless steel bent spoon, padded handle 3	
113		Plastic spoon with long handle (for babies) Long handles bright color spoon	
114		Plastic glass with rim cut on one side Plastic glass with one side of the rim is cut to accommodate nose:3	
115		Stainless steel plates with high rim: High rim to prevent spilling over of food	
116		Spouted cups	
117	Dental Equipment	Portable dental darkroom along with consumable Developer by Fixer	

### 9. List of Physiotherapy Equipment to be Purchased for DEIC

Sl no	Name of the Equipment	Remarks
1	Activity Table/ Modified Wheel chair (Paediatric)	
2	Rowing	
3	Parralal Bars (Paediatric)	
4	Weight Cuffs (Paediatric)	
5	Weight & Puller Exercise (Paediatric)	
6	Ankle Exerciser (Paediatric)	
7	Shoulder Wheel (Paediatric)	
8	Fore - Arm- Exerciser (Paediatric)	
9	Tread Mill (Paediatric)	
10	Tilt Table (Paediatric)	
11	Multi Gym (Paediatric)	
12	Wrist Roll (Paediatric)	
13	Grip Exerciser (Paediatric)	
14	Floor Ladder (Paediatric)	
15	Wrist Circumductor (Paediatric)	
16	Knee Hammer (Paediatric)	
17	Depth Perception Peg Board	
18	Treatment Cuff (6x4x2, 6x4x1)	
19	Height adjustable walker and rollator	
20	Standing frame (Paediatric)	
21	Nimals Hand Table apparatus (Paediatric)	
22	CP Chair (Paediatric)	
23	Corner Seat (Paediatric)	
24	Crawler (Paediatric)	
25	Staircase (Paediatric)	
26	Beach Ball with different sizes (Paediatric)	
27	Peanut Ball with different sizes	
28	Electrical stimulator	
29	Therapeutic Ultrasound Machine (Paediatric, Double Headed)	

### 10. List of Equipment's and Furniture including Reagents and Diagnostic Kits for UPHC

Sl No	Name of the EIF	Remarks
1	Equipment for New Born Care and Neonatal	
2	IUCD insertion kit	
3	Refrigerator	
4	Waste disposal twin bucket, hypochlorite solution/bleach	
5	Binocular microscope	
6	Adult weighing scale	
7	Baby weighing scale	
8	Height measuring Scale	
9	Sponge holding forceps	
10	Vulsellum uterine forceps	
11	Tenaculum uterine forceps	
12	Torch without batteries	
13	Bowl for antiseptic solution for soaking cotton swabs	
14	Kits for testing residual chlorine in drinking water	
15	H <sub>2</sub> S Strip test bottles.	
16	Head Light	
17	Ear specula.	
18	B.P. Apparatus table model	
19	Stethoscope	
20	Artery Forceps	
21	Eles forceps	
22	Needle Holder	
23	Cutting needles of different Sizes- Big	
24	Cutting needles of different Sizes- Medium	
25	Cutting needles of different Sizes- Small	
26	Suturing materials silk	
27	Scissors	
28	Stitch Cutter	
29	Toothed forceps	
30	Plain forceps	
31	Tooth extractor Universal	
32	Mouth gag	
33	Gastric tube for waxing	
34	Needle Cutter of different Sizes	
<b>Desirable</b>		
35	Ear Syringe	
36	Otoscope	
37	Nebuliser	
<b>List of equipment for Pap smear</b>		
38	Cusco's vaginal speculum Small Size	
	Cusco's vaginal speculum Medium Size	
	Cusco's vaginal speculum Large size	

Sim's vaginal speculum – single (Small Size)	
Sim's vaginal speculum – single (Medium Size)	
Sim's vaginal speculum – single (Large Size)	
Sim's vaginal speculum – double ended -(Small size)	
Sim's vaginal speculum – double ended -(medium Size)	
Sim's vaginal speculum – double ended -(large size)	
Sterile Gloves	
Sterilised cotton swabs and swab sticks in a jar with lid	
Kidney tray for keeping used instruments	
Bowl for antiseptic solution	
Antiseptic solution: Chlorhexidine 1% or Cetrimide 2% (if povidone iodine solution is available, it is preferable to use that)	
Cheatle's forceps	
For vaginal and Pap Smears:	
Clean slides with cover slips	
Cotton swab sticks	
KOH solution in bottle with dropper	
Saline in bottle with dropper	
Ayre's spatula	
Fixing solution/hair spray	
<b>Requirements of the laboratory</b>	
<b>Essential Reagents</b>	
39	Uristix for urine albumin and sugar analysis
40	ABO & Rh antibodies
41	Gram's iodine
42	Crystal Violet stain
43	PH test strips
44	RPR test kits for syphilis
45	H2S Strip test kits for faecal contamination of drinking water
46	Test kits for estimation of residual chlorine in drinking water using orthotoludine reagent
47	1000 Nos Whole Blood Finger Prick HIV Rapid Test and STI Screening Test each
48	1% hypo chloride solution
<b>Glassware and other equipment:</b>	
49	Colorimeter
50	Test tubes
51	Pipettes
52	Glass rods
53	Glass slides
54	Cover slips
55	Differential blood cell counter (Desirable)
56	Glucometer (Desirable)
<b>List of Furniture &amp; Fixtures</b>	

57	Examination table	
58	Wheel chair	
59	Stretcher on trolley	
60	Foot step	
61	Stool	
62	Medicine chest	
63	Buckets	
64	Dustbin (BMW Buckets)	
<b>Essential Medical/Surgical items</b>		
65	Blood Pressure Apparatus (Non-mercury is desirable)	
66	Stethoscope	
67	Tongue Depressor	
68	Thermometer Clinical	
69	Hub cutter	
70	Needle Destroyer	
71	Instrument trolley	
72	I V stand	
73	Macintosh for labour and OT table	
74	One observation bed	
75	Mattress	
76	Bed sheet	
77	Pillows	

### 11. EQUIPMENT, CONSUMABLES AND MISCELLANEOUS SUPPLIES FOR HWC.

Type	Items		
<b>I. Items required for routine services</b>			
Furniture and Fixtures	Chairs for patient waiting area		
	Bedside screen (3 fold) in LR & wards		
	Steel Almirah / Cupboard/storage chests		
	Examination table		
Clinical Material, tools and equipment	Basin 825 ml. Ss (Stainless Steel)		
	Basin deep (capacity 6 litre)		
	Instrument tray with cover 310 x 195x63mm SS		
	Dressing Drum with cover 0.945 liters stainless steel		
	Hemoglobinometer (strip/micro-cuvette)		
	Weighing Scale, Adult 125 kg/280 lb		
	Instrument sterilizer		
	Sphygmomanometer Aneroid 300 mm with cuff		
	Kelly's hemostat Forceps straight 140 mm		
	Cheate's Forceps		
	Cheate holder		
	Needle holder		
	Scalpel holder with blade		
	Sponge holding forceps		

Type	Items		
	Kidney tray		
	Artery Forceps, straight, 160mm Stainless steel		
	Sitch cutter		
	Alles forceps		
	Dressing Forceps (spring type), 160 mm, stainless steel		
	Stethoscope		
	Hub Cutter/Needle Destroyer		
	I/V Stand		
	Artery Forceps-Curved		
	Digital Thermometer		
	Examination Lamp		
	Oxygen Cylinder with trolley		
	Stadiometer		
	Nebulizer		
Gauze Cutting Scissors Straight			
Lab - Diagnostic Materials	Slide drying rack		
	Specimen collection bottle		
	Micro-pipette		
	Yellow Tips for Micropipette		
	Microscope		
	Neuber Chamber		
	ESR Stand with tube		
	TC-DC Count apparatus		
	Stopwatch		
	Sickeling test kit		
	Water quality test- H2S Strip test		
Colorimeter			
Centrifuge machine			
<b>2. Items required for providing RMNCHA services</b>			
Clinical Material, tools and equipment	Weighting Scale (10 Kg)		
	Vulsellum Uterine Forceps curved 25.5 cm		
	Cusco's/Graves Speculum vaginal bi-valve		
	Sims retractor/depressor		
	Sims Speculum vaginal double ended ISS Medium		
	Uterine Sound Graduated		
	Cord cutting Scissors, Blunt, curved on flat, 160 mm ss		
	Foetoscope		
	Ambu Bag (Paediatric size) with Baby mask		
	Episiotomy Scissors		
Lab Materials	Rapid Pregnancy Testing Kit		
	Dipsticks for urine test for protein and sugar		
<b>3. Items required for providing Speciality services</b>			
Oral Health	Interdental cleaning aids		
	Dental Probe		
	Universal tooth extractor		
ENT	Flashlight/Torch Box-type pre-focused (4 cell)		
	Tongue Depressor		
	Mouth Gag		
	Mouth Mirror		
	Tuning fork		
	Nasal Speculum (St. Claire's)*		

Type	Items		
	Ear Speculum – metallic, dull finish		
	Jobson-Horne probe		
	Otoscope		
Communicable diseases	Malaria Rapid Test Kit		
	Whole blood finger prick HIV Rapid Test & STI screening test		
	Typhoid rapid test kit		
	Rapid test kit for Hepatitis B & C		
NCD	Glucometer with strips and lancet		
	VIA Kit for screening cervical cancer		
Ophthalmic services	Snellen vision chart		
	Near vision chart		

**Annexure-1**  
**Technical Specification:**

**Baby Weighing Scale with Sling**

**Technical Specification:**

- Should have Tough Nylon Plastic (Black) body with demonstration photos in white colour
- Capacity : Maximum load capacity should be 5kg
- Each product should be stamped by weight & measurement department.
- Should have color coded reading scale.
- Should have Zero adjustment.
- Scale should be should have inbuilt overload protection.
- Minimum graduation should not less than 500 gms.
- Scale should be fitted with an approx.3" rod handles with non-corrosive metallic ring on top and S shaped hook at the bottom for suspending the sling bag.
- Spring used in the scale should be made of heavy duty steel spring, non-corrosive grade-II.
- Sling bag should be made of double stitched tough & tear resistant 100% polyester to weigh the baby (With load carrying capacity of at least 7 kgs.) Sling bag width at the middle of the bag: 24". Two nos. Sling bags should be provided in each baby weighing scale.

**Clinical Digital Thermometer**

**Technical Specification:**

- Digital thermometer with Fahrenheit scale.
- Safe to use, automatic, no glass, no mercury.
- Measurement accuracy: +/- 0.2 °F between 98.0 °F to 102.0 °F

+/- 0.3 °F between 96.4 °F to 98.0 °F  
 +/- 0.3 °F between 102.0 °F to 106.0 °F

(Type testing certificate from lab relating to measurement accuracy need to furnished in technical bid)

- Should have auto shut off facility
- Display should be LCD type
- Beeper alarm of approx. 10 seconds sound signal when peak temperature reached.
- Average measurement time: Oral: 30 sec, Rectal: 30sec, Auxiliary: 60 sec.
- Battery operated
- Should have more than 200hrs of battery life.
- Warranty: 1 year

### Technical specifications for Equipment/ Instruments for HWCs.

#### 1. Furniture and Fixtures

##### Quality Standard:

- Should be CE/BIFMA/BIS approved model.
- Manufacturer should have ISO 9001 certification for quality management standards.
- Manufacturer should have ISO 14001 certification for environmental management systems.
- Manufacturer should have OHSAS 18001 certification for occupational health & safety management.
- Should furnish stainless steel grade certificate from Govt/Govt. approved testing laboratory.
- Manufacture should produce test certificate from Govt/Govt. approved laboratory for test procedure like impact test, bend test, salt spray chamber test, epoxy powder coating & phosphate coating for quoted item

Items	Specifications
Chairs for patient waiting area	The chair should be heavy duty construction of three seater capacity. The frame structure shall be constructed with durable steel able to resist to rust. All joints shall be welded and all exposed surfaces shall be coated with a layer of anti-corrosion paint follow by another 2 finishing coats either powder coating or nylon polyester coating to withstand periodic cleaning and typical abuse. Color: Black and Silver Size: (L)1700mm*W550mm*480mm Approximate 10mm +/- Material: Metal
Bed sight Screen (Screen Separators with stand)	<ul style="list-style-type: none"> <li>• 3 folding partitions - 6 feet high and 6 feet long when opened</li> <li>• Tubular frame (CRCA tubular) mounted on 5cms high quality corrosion free castors with fine quality curtains.</li> <li>• Pre-Treated and Epoxy Powder coated.</li> </ul>
Steel Almirah /	<ul style="list-style-type: none"> <li>• Product Size: 900 mm (W) x 480mm (D) x 1900 mm (H) Height</li> </ul>

Items	Specifications
Cupboard/storage chests	<p>(Approximate 20mm +/-)</p> <ul style="list-style-type: none"> <li>• Construction &amp; Material: Minimum 0.8 mm thick CRCA for Shelf &amp; 0.9 mm thick CRCA for all other components.</li> <li>• Locking: 2 way locking mechanism with shooting bolt provided chrome plated with Handle</li> <li>• Shelving: Minimum 4 nos with uniformly distributed load capacity per full shelf is 40-50 Kg maximum Finished with Epoxy Polyester Powder coated.</li> </ul>
Examination Table	<p><b>Technical Specification</b></p> <ul style="list-style-type: none"> <li>• Overall approx size: 1820 mm L x 570mm W x 820mm H (L x W x H) ±20mm</li> <li>• Two section with two fold Cushion top of 3"mm thickness made of 40density PU foam</li> <li>• Tubular construction, machine pressed double bend mild steel sheet</li> <li>• Headrest adjusted on ratchet</li> <li>• Legs fixed with rubber feet.</li> <li>• Pre-treated and epoxy powder coated</li> <li>• Patient load bearing capacity of 120 kgs minimum.</li> </ul> <p><b>Foot Step</b></p> <p>The step stools shall have a maximum load capacity of 150 kg.</p> <p>Stainless steel frame shall be made of square /round steel tubing 20 x 20 mm.</p> <p>Should be double-step type with step heights of are 200 and 400 mm (+/-10mm) respectively.</p> <p>It should be provided with slip-resistant steps are made of black moulded polyurethane (PU) coated foam plates of minimum 15mm thickness.</p> <p>Overall dimension: 590x410x200/400 mm.</p>
I/V Stand	<p>Overall approx. size: height – 150cm-230 cm (with telescopic adjustable height)</p> <p>Main Frame: Strong &amp; Sturdy stainless steel tubular construction mounted on four pronged tubular/rectangular base fitted with five swivel rust proof castors of 50mm diameter.</p> <p>Stainless steel rod with double hooks</p> <p>All the Stainless Steel should be of 304 grade</p> <p>Should be pre-treated and epoxy coated finish.</p>
<b>Clinical Material, tools and equipment</b>	

Items	Specifications
Basin 825 ml. Ss (Stainless Steel)	Basin of 825 ml. Ss (Stainless Steel)
Basin deep (capacity 6 litre)	SS Basin square/rectangular shaped of capacity 6liter.
Tray instrument/Dressing with cover 310 x 195x63mm SS.	Instrument Tray with Cover Made of 304 grade Stainless Steel Size: 310 x 195x63mm Manufacturer should be ISO13485 approved Product should be BIS/CE approved
Dressing Drum with cover 0.945 litres stainless steel	<ul style="list-style-type: none"> <li>• Should be made of joint-less stainless steel of 304 grade steel of 0.5mm thickness.</li> <li>• Should have perforated body.</li> <li>• Should have chain locking with clamp to open or close the perforated body.</li> <li>• Size : 0.945 litres</li> <li>• Manufacturer should be ISO13485 approved</li> <li>• Product should be BIS/CE approved</li> </ul>
Hemoglobinometer (strip/micro-cuvette)	<p>It should be a point of care instrument to measure hemoglobin. The unit should have LCD screen for instant display of test result.</p> <ul style="list-style-type: none"> <li>• It should be Factory calibrated as per ICSH (International Council for Standardization in Hematology). Needs no further calibration</li> <li>• Self-checking facility should be there between every measurement</li> <li>• Shelf life of the test strip/cuvette should be minimum 90 days from the opening of the packet/ box and at least one year from the date of manufacturing.</li> <li>• Measurement Range: 4gm./dL to 20gm./dL</li> <li>• Accuracy should be <math>\geq 96\%CV &lt; 5\%</math></li> <li>• Sample volume; approximately 10 to 15 <math>\mu</math>L whole blood; capillary or venous blood</li> <li>• Minimum stored result memory :250 results</li> </ul>

Items	Specifications
	<ul style="list-style-type: none"> <li>• USB interface facility should be there for transfer of stored results.</li> </ul> <p><b>Soft keys for user operation.</b> It should communicate through mini-USB connector with a Computer/ laptop.</p> <p>Battery operated; The battery should support at least 500 tests.</p> <p>Each device should supplied with: 1. Cover (Wash Proof) 2. USB Cable for data transfer 3. Other items required for functioning of the machine. The cost of Device, Strips (50 Nos.) &amp; lancets (55 Nos.) should be quoted separately in price BOQ excel format.</p> <p>Operating condition: Capable of operating continuously in ambient temperature of 0 to 50 deg. C and relative humidity of 15 to 90% in ideal circumstances.</p> <p><b>Safety &amp; Quality Standard:</b> European CE or USFDA Certified. It should be Factory calibrated as per ICSH (International Council for Standardization in Hematology). The manufacturer should be ISO 13485 certified.</p>
Weighing Scale, Adult 125 kg/280 lb	<p>Sturdy <b>dial type mechanical</b> platform weighing machine for adult and children.</p> <p>Zero adjustment facility should be there.</p> <p>Sensitivity: 500 gm</p> <p>Range of weighing: 0-125kg</p> <p>The manufacturer shall have the valid manufacturing license and should have model approval by the legal metrological Deptt. and the weighing scale must be stamped by the by legal metrological Deptt. In case of distributor, the bidder should have valid distributor and repair license from legal metrological Deptt. Govt., of Odisha.</p> <p><b>ISO. Certified manufacturer</b> (certificate to be submitted), Warranty: 1 Year</p>
Instrument Sterilizer	<p>Electrically operated.</p> <p>Size ideally: 24x8x6 inch</p> <p>Should have replaceable immersion type heater ISI marked.</p> <p>Body should be made of SS-304 grade steel. (Test report certificate from independent laboratory should be submitted with technical bid.)</p> <p>Power Supply</p> <p>Power input to be 220-240VAC, 50Hz fitted with Indian plug</p> <p>Manufacturer should be ISO13485 approved</p> <p>Product should be BIS/CE approved</p>
Sphygmomanometer Aneroid 300 mm with cuff	<p><b>Technical Specification of BP Apparatus(Aneroid)</b></p> <ul style="list-style-type: none"> <li>• Corrosion resistant shock proof body, chrome plated metal/ stainless steel pressure control valve, scale 0-300 mm hg with accuracy of +/- 3mmHg</li> <li>• Air release at closed lap with maximum 4mmHg/Minute.</li> <li>• Manual setting of deflation possible upto 2/3mm Hg/sec. From</li> </ul>

Items	Specifications
	<p>260mmHg to 15mm Hg in a maximum deflation time of 10 seconds.</p> <ul style="list-style-type: none"> <li>• Gauge's background in white colour.</li> <li>• Graduated scale for ever/ 2mmhg, every 10 units and every 20units.</li> <li>• Nylon straps cuff with pouch, latex bulb with completely chromium plated valve with regulation of vent-hole air by screw valve.</li> <li>• The cuff is inflated just to fit in the limb for which an inflation bulb is used to control the air pressure within the cuff.</li> <li>• The rubber tubes used should have an internal diameter of <math>3 \pm 0.5</math>mm and the external diameter should not be less than 8mm; The dial manometer with diameter of 50 mm-60mm</li> </ul> <p><b>Accessories (mandatory, standard, optional)</b>-adult arm cuffs of size medium &amp; large and pediatric size, inflation bulb, tubing</p> <p><b>STANDARDS AND SAFETY</b>  Manufacturer should be ISO13485; should be USFDA, CE, BIS or UL approved product</p>
Surgical Instruments	<p><b>Quality standard (Applicable to all surgical instruments):</b>  Instruments should be made up of stainless steel medical grade AISI 410 &amp; 420. Test reports should be submitted in the technical bid. Part No. and the CE Mark must be engraved/Embossed on the instrument.  Sample demonstration  Instrument should cover 2 years replacement warrantee.  All the instruments should be autoclavable</p>
Kelly's Forceps	Kelly's hemostat Forceps straight 140 mm
Cheate's Forceps with Holder	Cheate's Forceps 10" with SS Holder
Needle Holder	Needle holder -8inch
Scalpel holder with blade	Central Store supply
Sponge holder	Sponge holding forceps 8" with double action Jaws
Kidney tray	Kidney tray, 8"
Artery Forceps, straight, 160mm Stainless steel	Artery Forceps, straight, 160mm Stainless steel
Stitch cutter	Stitch cutting size with round handle double action jaw of 6 inch.
Allis forceps	Tissue Forceps finger ring, ratcheted forceps straight with 5x6 teeth and an overall length of 7-1/2 inches.
Dressing Forceps (spring type), 160 mm, stainless steel	Dressing Forceps (spring type), 160 mm, stainless steel
Stethoscope	Stethoscope of standard size, chromium plated metal binaural, V rubber tube in one piece. Rotating piper fitting for both flip functions. Doubleheadpaediatric&Adult stethoscope.

Items	Specifications
	<ul style="list-style-type: none"> <li>• Extra-soft, replaceable and pivot able ear-tips for perfect sealing at the ear canal.</li> <li>• Designed with Precision chest-piece made of stainless steel/ chromed brass.</li> <li>• Good quality diaphragm of minimum -Ø-45mm</li> <li>• High quality membrane for precise acoustics with non-chill rims for improved adaptation on the skin and for excellent sound transmission.</li> <li>• Length should be 27" to 29 with preferable colour -black.</li> <li>• The Y-tube should be made of Latex-free treated rubber.</li> </ul> <p>Easy to dismantle, and therefore to clean and disinfect.  Manufacturer should be ISO13485 approved  Product should be BIS/CE approved</p>
Hub Cutter and Needle Destroyer	<ul style="list-style-type: none"> <li>• Should be lightweight, portable &amp; compact.</li> <li>• Housing should be moulded type, shock proof and made of ABS plastic/Stainless steel of 304 Grade.</li> <li>• Should be provided with removable discharge tray for easy disposal of syringe hubs.</li> <li>• Should have provision to burn the needle and cut the syringe tip.</li> <li>• Should have high Carbon Steel cutter to cut the syringe tips.</li> <li>• Should able to cut and destroy the needle up to 18G.</li> <li>• Should able to destroy minimum of 5 injection needles on continuous operation.</li> <li>• Should have heavy duty transformer.</li> <li>• Should have power ON/OFF switch with visual indication.</li> <li>• Should be properly insulated for protection from electric hazards.</li> <li>• Should have fuse protection with 5no.of fuse to be supplied of adequate rating.</li> </ul> <p><b>Power Supply:</b>  Power supply should be 220-240 V AC, 50Hz with Indian plug.  Manufacturer should be ISO13485 approved  Product should be BIS/CE approved</p>
Artery Curved	Forceps- Artery Forceps 5"Curved double action jaw ratcheted with ring handle double action jaws.
Gauze Scissors	Cutting Angled with the lower blade being slightly longer & tip of the lower blade features a flattened blunt nodule to slide between bandages and skin without harming the skin. Size:5"
Digital Thermometer	Range of temperature measurement 320C- 420 (89.60F-109.40F). Can be calibrated in both centigrade and Fahrenheit, but if only one option is available, then Fahrenheit is preferable.

Items	Specifications
	Buzzer signal function. Takes 10-15 seconds to measure temperature. Can be used in the armpit/axilla, orally and rectally. Accuracy of temperature $\pm 0.1$ degC and $\pm 0.2$ F. User's interface :LCD display Manufacturer should be ISO13485 approved Product should be BIS/CE approved
Examination Lamp	<ul style="list-style-type: none"> <li>• Source of Light: LED light</li> <li>• Illumination(Ix) should be LED</li> <li>• Minimum 25,000 Lux at a working distance of .5m</li> <li>• Radial and axial movement of the lamp</li> <li>• Colour temperature: minimum 4000K</li> <li>• Minimum Diameter of the spot 5cm</li> <li>• Flexible Light pipe for easy positioning of minimum length of 1meter.</li> <li>• Lamp Life of minimum 30000 hrs</li> <li>• CRI&gt;90</li> <li>• Power Requirements-Recharging unit: Input voltage- 220V-240V AC, 50Hz</li> </ul> Should be USFDA/CE approved model. Manufacturer should be ISO 13485 certified. Electrical safety conforms to the standards for electrical safety IEC 60601-1General requirements (or equivalent BIS Standard)
Oxygen Cylinder with trolley	Central Store Supply
Stadiometer	<ul style="list-style-type: none"> <li>• The measuring rod can be dismantled into several pieces and can be set easily.</li> <li>• The scale must be printed along the side of the measuring rod.</li> <li>• Measuring range (Both in cm &amp; inch): 20-205 cm and 8 - 81".</li> <li>• Graduation of measuring rod: 1mm / 8inch.</li> <li>• The structure should be made of ABS plastic. The product</li> <li>• should be CE certified (certificate to be submitted in technical bid)</li> </ul> Warranty : 1 Year
Nebulizer	Should be of Heavy duty compact Nebulizer Heavy duty ,Compact, lightweight, low noise(50dB+ 3dB)

Items	Specifications
	<p>Durable long life compressor. Suitable for heavy duty/ institutional (hospital) use, should be able to run uninterruptedly for one hour.            Max Pressure = 2.0-2.5 bars            Operating pressure: 1 to 1.5 bars            Compressor Air flow: 8Lpm            Normal Air Flow: 4lpm            Should produce particle of size 1-5 micron.            Mass median Diameter (MMD) ~2.5-3m.            Output rate: 500gm/Min.            Made of Heavy duty ABS body            Power Supply: Power input to be 220-240VAC, 50Hz fitted with Indian plug            Should be USFDA/CE approved model.            Manufacturer should be ISO 13485 certified.            Electrical safety conforms to the standards for electrical safety IEC 60601-1 General requirements (or equivalent BIS Standard)</p>
<b>Lab -Diagnostic Materials</b>	
Slide drying rack	<p>Acrylic slide rack, inclined, used for draining/drying slides.            5 slots accommodate 25 slides.</p>
Specimen collection bottle(100/pack)	Central Store
Micropipette	<ul style="list-style-type: none"> <li>• Should have ergonomic design with light &amp; smooth plunger action</li> <li>• Should have soft feel handle grip having both left &amp; right hand operation</li> <li>• Pipette handle should have thermoplastic elastomeric to prevent transfer of body heat to pipette volume during continuous usage</li> <li>• Fully autoclavable: Entire pipette can be steam autoclaved at temp. Of 121 0C</li> <li>• Should have larger &amp; clear 3 digit display giving smaller increment for wider selection of volume</li> <li>• Volume range should be of 1-10ml with increment of 20 µl</li> <li>• Accuracy: 1 to 2%</li> <li>• Should have locking mechanism to prevent accidental volume change during pipetting</li> <li>• Should have one hand eject facility</li> <li>• Should have in house clinical, repair and calibration facility</li> <li>• The tip cone should have leak free operation, smooth and light loading operation with choice of using variety of tips.</li> <li>• Should be compatibles to universal tip types</li> <li>• Should be available with different color codes.</li> <li>• Warranty: 3years.</li> </ul>

Items	Specifications
	<p><b>Quality Standards:</b></p> <ul style="list-style-type: none"> <li>• Should be USFDA/CE (IVD) approved product</li> <li>• Manufacturers should have ISO 13485 certification for quality standards</li> <li>• Should be applied with individual QC &amp; calibration report according to ISO 8655</li> </ul>
Yellow Tips for Micropipette	<p>Pipette tips should be designed to fit and function on a wide variety of single or multichannel pipettes. It should be manufactured from the finest grade polypropylene material for proper fit and straightness. Packaging: Sterile wrapped racks.</p> <p><b>Quality Standards:</b> The item should be CE certified Manufacturer should be ISO 13485 Certified</p>
Microscope	<p><b>GMDN name :</b> Binocular Microscope</p> <p><b>A. General Use</b></p> <ul style="list-style-type: none"> <li>• <b>Clinical purpose:</b> Binocular microscope is simply a microscope that lets the viewer use both eyes. The microscope has 2 eye lenses. The development of the double eye piece microscope was adapted to reduce the eyestrain and muscular strain that typically results from traditional microscopes.</li> <li>• <b>Used by clinical department/ward:</b> Clinical labs</li> </ul> <p><b>B. Technical Characteristics</b></p> <ul style="list-style-type: none"> <li>• <b>Technical characteristics (specific to this type of device)</b> <ul style="list-style-type: none"> <li>• Body-Single mould sturdy stand, inclined Binocular body 30 °, 360° rotatable head</li> <li>• Eyepieces-Highest quality 10X/20mm wide angle anti fungus field eyepiece. One with pointer. Diopter adjustment must be present on both eye pieces.</li> <li>• Objectives-Parfocal, antifungus coated 4x, 10x, 40x and 100x (oil immersion) with semi planner achromatic correction. Objective should be well centred even if their position on turret is changed.</li> <li>• Optical system-Infinity corrected</li> <li>• Stage - Double plate rackless horizontal mechanical stage preferably 100 x 140 mm with fine vernier graduations designed with convenient coaxial adjustment for slide manipulation preferably through 30 x 70 mm double slide holder</li> <li>• Sub stage-Abbe condenser focusable, continuously variable iris diaphragm</li> <li>• Illuminator-Built-in LED light source with white light with intensity control and LED life of more than 10,000 Hrs. Colour Temperature minimum: 4000K.</li> <li>• Finish-A durable textured acid resistant finish.</li> </ul> </li> </ul>

Items	Specifications
	<ul style="list-style-type: none"> <li>• Battery backup: minimum 1 Hour</li> <li>• Nose piece: Backward tilted revolving nose piece suitable to accommodate four objectives with click stop and rubber grip.</li> <li>• Focussing: Coaxial coarse and fine focussing knob, capable of smooth, fine focussing movement sensitivity; minimum: 300 micron; focussing stop for slide safety</li> <li>• 12-Objectives: All DIN type</li> </ul> <p>Plan Achromatic (Anti Fungus)</p> <ul style="list-style-type: none"> <li>• 04x N.A .01 WD 6.50mm</li> <li>• 10x N.A 0.25 WD 5.6mm</li> <li>• 40x N.A 0.65 WD .6mm</li> <li>• 100x (Oil).N.A 1.25 WD .13mm</li> </ul> <ul style="list-style-type: none"> <li>• User's interface : Manual</li> <li>• Software and/or standard of communication (where ever required) : NA</li> </ul> <p><b>C. Physical Characteristics</b></p> <ul style="list-style-type: none"> <li>• Dimensions (metric) : NA</li> <li>• Weight (lbs, kg) : NA</li> <li>• Capacity: NA</li> <li>• Noise (in dBA) : NA</li> <li>• Heat dissipation: NA</li> <li>• Mobility, portability: Portable</li> </ul> <p><b>D. ENERGY SOURCE :( Electricity)</b></p> <ul style="list-style-type: none"> <li>• Power Requirements: Input voltage- single</li> <li>• Battery operated: Yes with 1hour backup</li> <li>• Tolerance (to variations, shutdowns): NA</li> <li>• Pressure gauge : NA</li> <li>• Operating pressure: NA</li> <li>• Sterilizing pressure: NA</li> <li>• Protection: Should have over-charging cut-off with visual symbol.</li> <li>• Power consumption : less than 2 Watt</li> </ul> <p><b>E. Accessories, Spare Parts, Consumables</b></p> <ul style="list-style-type: none"> <li>• Accessories (mandatory, standard, optional); Spare parts (main ones); Consumables / reagents (open, closed system) : Should provide with wooden storage box, dust cover, immersion oil.</li> </ul> <p><b>F. Environmental and Departmental Considerations</b></p> <ul style="list-style-type: none"> <li>• Atmosphere / Ambiance (air conditioning, humidity, dust ...)</li> </ul> <p><b>Operating condition:</b> Capable of operating continuously in ambient temperature of 10 to 50 deg C and relative humidity of 15 to 90% in ideal circumstances.</p> <p><b>Storage condition:</b> Capable of being stored continuously in ambient temperature of 0 to 50 deg C and relative humidity of 15 to 90%."</p> <ul style="list-style-type: none"> <li>• User's care, Cleaning, Disinfection &amp; Sterility issues: <ul style="list-style-type: none"> <li>• Disinfection: Parts of the Device that are designed to come into contact with the patient or the operator should either be capable of easy disinfection or be protected by a single</li> </ul> </li> </ul>

Items	Specifications
	<p>use/disposable cover.</p> <ul style="list-style-type: none"> <li>▪ Sterilization not required."</li> </ul> <p><b>G. Standards and Safety</b></p> <ul style="list-style-type: none"> <li>▪ <b>Certificates (pre-market, sanitary, ..):</b> Performance and safety standards (specific to the device type);Local and/or international</li> <li>▪ Should be US FDA/CE (from a Notified body)/BIS approved product.</li> <li>▪ Manufacturer should have ISO 13485 certification for quality standards.</li> <li>▪ Electrical safety conforms to the standards for electrical safety IEC 60601-General requirements (or equivalent BIS Standard)</li> <li>▪ Certified to be compliant with IEC 61010-1, IEC 61010-2-40 for safety.</li> </ul> <p><b>H. Training and Installation</b></p> <ul style="list-style-type: none"> <li>▪ <b>Pre-installation requirements: nature, values, quality, tolerance</b> <ul style="list-style-type: none"> <li>▪ Availability of 5 amp socket;</li> <li>▪ Safety and operation check before handover;</li> </ul> </li> <li>▪ <b>Requirements for sign-off :</b> Certificate of calibration and inspection from the manufacturer</li> <li>▪ <b>Training of staff (medical, paramedical, technicians)</b> <ul style="list-style-type: none"> <li>▪ Training of users on operation and basic maintenance;</li> <li>▪ Advanced maintenance tasks required shall be documented</li> </ul> </li> </ul> <p><b>I. Warranty and Maintenance</b></p> <ul style="list-style-type: none"> <li>▪ <b>Warranty:</b> 3 years</li> <li>▪ <b>Maintenance tasks:</b> CMC 5 years, 2 PM Visits Annually.</li> </ul> <p>All Breakdown calls to be attended within 24 hrs of registration."</p> <ul style="list-style-type: none"> <li>▪ <b>Service contract clauses, including prices :</b> The spare price list of all spares and accessories (including minor) required for maintenance and repairs in future after guarantee / warranty period should be attached;</li> </ul> <p><b>J. Documentation</b></p> <ul style="list-style-type: none"> <li>▪ <b>Operating manuals, service manuals, other manuals</b> "Should provide 2 sets(hardcopy and soft-copy) of:- <ul style="list-style-type: none"> <li>▪ User, technical and maintenance manuals to be supplied in English/Hindi language along with machine diagrams;</li> <li>▪ List of equipment and procedures required for local calibration and routine maintenance;</li> <li>▪ Service and operation manuals (original and copy) to be provided;</li> <li>▪ Advanced maintenance tasks documentation;</li> <li>▪ Certificate of calibration and inspection"</li> </ul> </li> <li>▪ <b>Other accompanying documents:</b> List of important spares and accessories, with their part numbers and cost;</li> </ul> <p><b>K. Notes</b></p> <ul style="list-style-type: none"> <li>▪ Service Support Contact details (Hierarchy Wise; including a toll free/landline number) "Contact details of manufacturer, supplier and local service agent to be provided;</li> </ul> <p>Any Contract (AMC/CMC/add-hoc) to be declared by the</p>

Items	Specifications
	manufacturer;" <ul style="list-style-type: none"> <li>• Recommendations or warnings Any warning signs would be adequately displayed</li> </ul>
Neuber Chamber	<p>The counting grid is of 3mmx 3mm in size.</p> <ul style="list-style-type: none"> <li>• The grid should have 9 square subdivisions of width 1mm.</li> <li>• The central square should be split into 25 squares of width 0.2mm (200µm)</li> <li>• The haemocytometer should have two cover slips</li> <li>• The glass cover should be of squared shaped of width 22mm</li> <li>• Material: Thermal and shock resistant glass</li> <li>• The chamber should have two special clamps to avoid the cover glass to avoid edge-lift.</li> </ul> <p>Pipette</p> <p>The pipette should be made of glass of maximum capacity of 20,200 and 2000µl.</p> <p>Dimensions (metric): length-300mm +/-10mm, 2mm thickness.</p> <p>Autoclaving at 121degree centigrade</p> <p>Should be resistance to mechanical, chemical and over heating.</p> <p><b>STANDARDS AND SAFETY:</b></p> <ol style="list-style-type: none"> <li>1. Should be ISO 648/IS 117 approved product.</li> <li>2. Manufacturer and Supplier should have ISO13485/ ISO 9001 certification for quality standards.</li> <li>3.Certificate of calibration from the manufacturer</li> <li>4. The product should be USFDA/ CE marked</li> </ol>
ESR Stand with tube	<p><b>ESR Stand with Tubes</b></p> <ul style="list-style-type: none"> <li>➤ Screw type stand with cast iron base and highly nickel plated</li> <li>➤ The base shall have minimum 5tube holder</li> <li>➤ 2nos spare tubes should be provided with the ESR stand.</li> <li>➤ The witrobes tube should be made of <b>Soda-lime Glass/borociliate glass material.</b></li> <li>➤ <b>Dimension:</b> <ul style="list-style-type: none"> <li>• 110 to 120mm mm Length</li> <li>• 2.9mm to 3 mm inside Diameter</li> <li>• 7.0mm to 8.0mm outer diameter.</li> </ul> </li> <li>➤ The uniformity of the bore shall be <math>\pm 0.1</math> mm throughout the tube.</li> <li>➤ The tube shall be graduated from 0 to 105mm <math>\pm 0.25</math> mm in 1-mm divisions and numbered every 1 cm from the inside bottom of the tube.</li> <li>➤ The tube shall be legibly marked with the manufacturer's or vendor's name or mark and possess a frosted area for marking purposes.</li> <li>➤ Open at both ends; cotton plugged</li> <li>➤ Calibration in yellow/white for easy reading.</li> <li>➤ It should have good chemical resistance towards acids, salt solutions and organic solutions.</li> <li>➤ Can be easily cleaned with distle water.</li> <li>➤ It should be autoclavable.</li> </ul>
	<ul style="list-style-type: none"> <li>➤ The tube shall be legibly marked with the manufacturer's or vendor's name or mark and possess a frosted area for marking purposes.</li> <li>➤ Open at both ends; cotton plugged</li> <li>➤ Calibration in yellow/white for easy reading.</li> <li>➤ It should have good chemical resistance towards acids, salt solutions and organic solutions.</li> <li>➤ Can be easily cleaned with distle water.</li> <li>➤ It should be autoclavable.</li> </ul>

Items	Specifications
	<b>Quality Standards:</b> <ul style="list-style-type: none"> <li>➤ The item should be CE certified with marking</li> <li>➤ Manufacturer should be ISO 13485 Certified</li> </ul>
TC-DC apparatus	<b>TC-DC Count apparatus:</b> All metal case housing, durable, Chemical resistant easy to clean and disinfect. Stable and wide base with rubber feet Should be 8-Key model with dual reset mount Should have totalizer window Should have WBC maturation series picto-strip chart just above the key window
Stopwatch	Stopwatch for Laboratory use
Sickling test kit	Kit consists of R1-2x20ml(Solubility buffer) R2-2 vials (Solubility reagent-Sodium Dithionite) Empty reaction tube-20nos Result reading stand-1 Reagent dropper-2 Sample dropper-20 Rubber teat-2 Pack insert-1
Water quality test-H <sub>2</sub> S Strip test	<ul style="list-style-type: none"> <li>• Sensitivity: 5, 10, 20, 30, 40, 50, 60 and 80 ppm (mg/L)</li> <li>• Test time: 30 seconds</li> <li>• Bottle of 50 test strips</li> </ul>
Colorimeter	<b>Digital colorimeter for clinical biochemical test.</b> <b>Digital Colorimeter with auto zero facility displaying test result.</b> <b>Easy operation with auto storage facility of concentration.</b> Wavelength: 410, 470, 490, 520, 540, 580, 610 & 640 nm Photometric range :%of Transmission 0 to 100 Absorbance -0.0 to 1.5 Abs Repeatability ± 0.01 Abs Sample volume:1ml Light source: Halogen/LED Sample Holder :75 x 12 mm ± 0.5 mm Test tube with 10 mm Path Length With shutter. The manufacturer should be ISO13485 approved Product should be CE certified Product should be safety compliance to IEC 61010
Centrifuge machine	<b>Centrifuge Table Top- 4tube</b> GMDN name : <b>Centrifuge</b> 1      USE 1.1      Clinical purpose: Used in Biochemical and Analytical labs for Hematocrit, blood Corpusule percentage, Serum Analysis, Precipitate

Items	Specifications
	<p>Seperation and Blood Group matching.</p> <p>1.2 Used by clinical department/ward: Analytical Laboratories</p> <p><b>TECHNICAL</b></p> <p><b>2 TECHNICAL CHARACTERISTICS</b></p> <p>2.1 Technical characteristics (specific to this type of device) "</p> <ul style="list-style-type: none"> <li>• Speed: Maximum Range 3000 to 3500 RPM</li> <li>• Receptrocating Centrifugal force (RCF): 1500 to 1600</li> <li>• Minimum Capacity: 60ml</li> <li>• Digital Timer range: 0 to 30 minutes with Hold Function</li> <li>• Auto Lid interlock to prevent opening while running centrifuge with emergency lidlock release.</li> <li>• The lid will be only opened once the motor is completely stop.</li> <li>• Digital display of RPM</li> <li>• Stainless steel Chamber easy to clean</li> <li>• Hinges to prevent door falling</li> <li>• Rotor Head Sizes: 4x 15ml. Rotor Head: Angle rotor</li> <li>• Rotors should be autoclavable</li> <li>• The unit should be vibration free."</li> <li>• Noise level &lt;60dB.</li> </ul> <p>2.2 User's interface Manual</p> <p>2.3 Software and/or standard of communication (where ever required) NA</p> <p><b>3 PHYSICAL CHARACTERISTICS</b></p> <p>3.1 Dimensions (metric) NA</p> <p>3.2 Weight (lbs, kg) NA</p> <p>3.3 Capacity 120 ml or above</p> <p>3.4 Noise (in dBA) NA</p> <p>3.5 Heat dissipation Heat Dissipation: Should maintain nominal Temp and the heat should be disbursed through an cooling mechanism</p> <p>3.6 Mobility, portability Portable</p> <p><b>4 ENERGY SOURCE (electricity, UPS, solar, gas, water, CO2 ....)</b></p> <p>4.1 Power Requirements "220-240 V/50Hz</p> <p>4.2 Battery operated No</p> <p>4.3 Protection NA</p> <p>4.4 Power consumption 100 to 200 Watts</p> <p><b>5 ACCESSORIES, SPARE PARTS, CONSUMABLES</b></p> <p>5.1 Accessories (mandatory, standard, optional); Spare parts (main ones); Consumables / reagents (open, closed system) Rubber</p>

Items	Specifications
	<p>adapter should be provided for the use of vacutainer for 3ml and 5ml</p> <p><b>6 ENVIRONMENTAL AND DEPARTMENTAL CONSIDERATIONS</b></p> <p>6.1 Atmosphere / Ambiance (air conditioning, humidity, dust ...)</p> <p>1) Operating condition: Capable of operating continuously in ambient temperature of 10 to 50 deg C and relative humidity of 15 to 90% in ideal circumstances.</p> <p>2) Storage condition: Capable of being stored continuously in ambient temperature of 0 to 50 deg C and relative humidity of 15 to 90%.</p> <p>6.2 User's care, Cleaning, Disinfection &amp; Sterility issues</p> <p>"1) Disinfection: Parts of the Device that are designed to come into contact with the patient or the operator should either be capable of easy disinfection or be protected by a single use/disposable cover.</p> <p>2) Sterilization not required."</p> <p><b>7 STANDARDS AND SAFETY</b></p> <p>7.1 Certificates (pre-market, sanitary, ..); Performance and safety standards (specific to the device type); Local and/or international</p> <p>1. Should be FDA/CE (From a notified body) as per IVD /BIS approved product.</p> <p>2. Manufacturer and Supplier should have ISO 13485 certification for quality standards.</p> <p>3. Certified to be compliant with IEC 61010-1, IEC 61010-2-20 for safety, General requirements"</p> <p>IEC 61010-2-020 "Safety requirements for electrical equipment for measurement, control, and laboratory use - Part 2-020: Particular requirements for laboratory centrifuges"</p> <p><b>8 TRAINING AND INSTALLATION</b></p> <p>8.1 Pre-installation requirements: nature, values, quality, tolerance</p> <p>1) Availability of 5 amp socket;</p> <p>2) Safety and operation check before handover;</p> <p>8.2 Requirements for sign-off Certificate of calibration and inspection from the manufacturer</p> <p>8.3 Training of staff (medical, paramedical, technicians)</p> <p>1) Training of users on operation and basic maintenance;</p> <p>2) Advanced maintenance tasks required shall be documented</p> <p><b>9 WARRANTY AND MAINTENANCE</b></p> <p>9.1 Warranty 3 years</p> <p>9.2 Maintenance tasks "CMC 5 years 2 PM Visits Annually. All Breakdown calls to be attended within 24 hrs of registration."</p> <p>9.3 Service contract clauses, including prices The spare price list of all spares and accessories (including minor) required for maintenance and repairs in future after guarantee / warranty period should be attached;</p> <p><b>10 DOCUMENTATION</b></p> <p>10.1 Operating manuals, service manuals, other manuals "Should provide 2 sets (hardcopy and soft-copy) of:-</p> <p>1) User, technical and maintenance manuals to be supplied in English/Hindi language along with machine diagrams;</p>

Items	Specifications
	2) List of equipment and procedures required for local calibration and routine maintenance; 3) Service and operation manuals (original and copy) to be provided; 4) Advanced maintenance tasks documentation; 5) Certificate of calibration and inspection" 10.2 Other accompanying documents List of important spares and accessories, with their part numbers and cost; 11 NOTES 11.1 Service Support Contact details (Hierarchy Wise; including a toll free/landline number) "Contact details of manufacturer, supplier and local service agent to be provided; Any Contract (AMC/CMC/add-hoc) to be declared by the manufacturer;" 11.2 Recommendations or warnings any warning signs would be adequately displayed
<b>Items required for providing RMNCHA services</b>	
Weighing Scale, Infant (10 Kg)	<ul style="list-style-type: none"> <li>• Table top, light and portable,</li> <li>• Built in rechargeable battery,</li> <li>• Easy to clean baby tray (acrylic),</li> <li>• Zero weight adjustment facility,</li> <li>• Quick, clear digital read outs with LCD display</li> <li>• Measurement does not change with position of baby on the pan;</li> <li>• Provision to measure the height of the baby in its laying position.</li> <li>• Accuracy: 10g, resolution: 1g, limit: 10gm – 10kg</li> <li>• Settings-Auto setting to 0.00 once a the machine is switched on or when no external weight has been put on</li> </ul> <p><b>Dimension :Base: 300mm x 265mm x 85mm, Pan: 510mm x 300mm x 85mm/As per IS standard</b>  <b>Power Requirements-230 V AC &amp; Battery operated-6V, one hour backup</b>            The Scale should be as per BIS specifications.            The scale should have ISI mark ie IS: 2489 Or CE/USFDA certified.            Should have model approval from Legal Metrology Dept., Govt. of India.            The manufacturer shall have the valid license and should have model approval by thelegal metrological Deptt. And the weighing scale must be stamped by the by legalmetrologicalDeptt. In case of distributor, the bidder should have valid distributor andrepair license from legal metrological Deptt., Govt. of Odisha.</p>
Vulsellum Uterine	Vulsellum Uterine Forceps curved 25.5 cm

Items	Specifications
Cusco's/Graves Speculum vaginal bi-valve	Cusco's/Graves Speculum vaginal bi-valve small size
Sims retractor/depressor	Sims retractor/depressor small size
Sims Speculum vaginal double ended ISS Medium	Sims Speculum vaginal double ended ISS Medium size
Uterine Sound Graduated	Uterine Sound Graduated 10-12"
Cord cutting Scissors	Cord cutting Scissors- Size 6 inch double action jaw with round handle. Blunt, curved on flat, 160 mm ss
Foetal Doppler	<p>A portable, hand-held, battery-powered device assembly consisting of a measuring and display unit and an attached probe or interchangeable probes designed to noninvasively detect foetal heart beats using ultrasound/Doppler technology.</p> <ul style="list-style-type: none"> <li>• Water proof probes of 2MHz &amp; 3MHz frequency.</li> <li>• Ultra sound Intensity &lt;10mw/ cm<sup>2</sup>.</li> <li>• Auto Shut Of Facility to save Battery Power</li> <li>• Built -in Speaker, Volume Control Facility and Audio Output for Ear Phone</li> <li>• Heart Rate Range should be from 50 to 200 bpm with accuracy of + /-2%.</li> <li>• Should be Water Proof Body, Should have Facility for FHR Data transfer to PC.</li> <li>• LCD display</li> </ul> <p>Power Requirements AA batteries  Battery operated AA battery type; Minimum Battery Time of 300 minutes.</p> <p><b>Accessories:</b></p> <ul style="list-style-type: none"> <li>▪ AA battery(rechargeable)-2nos</li> <li>▪ 2MHz probe-1no</li> </ul> <p><b>STANDARDS AND SAFETY:</b></p> <ul style="list-style-type: none"> <li>▪ USFDA or CE(Notified) or UL approved product.</li> <li>▪ Type B or BF, Performance and safety standards (specific to the device type)</li> <li>▪ Shall meet IEC-60601-1-2:2007 Medical electrical equipment - Part 1-2: General requirements for basic safety and essential performance</li> <li>▪ Local and/or international Manufacturer should be ISO 13485 certified</li> </ul>
Suction Machine	<b>Technical Specification</b> 0-760 mm Hg ± 10 regulable

Items	Specifications
	<p>1/4 HP; single phase motor flutter free vacuum control knob Wide mouthed 1 LITRE (Polycarbonate) with self-sealing bungs and mechanical over flow safety device. Dimensions (metric) : Portable Table top Noise (in dBA) <b>50 dB A ± 3</b> Accessories &amp; Spares Collection container &amp; its cap, suction tube tips, a vacuum gauge and control knob.</p> <ul style="list-style-type: none"> <li>• Tubing: 8 mm ID x 2 mtr (PVC)</li> <li>• 1 lt polycarbonate jar of 1 nos</li> </ul> <p><b>Quality Standard:</b> Quoted model should be USFDA /CE certified.ISO 13485:2003; ISO 10079-1- 1999</p>
Ambu Bag (Paediatric size) with Baby mask	<ul style="list-style-type: none"> <li>• Manual resuscitator with transparent face-mask</li> <li>• Child models (750ml, &amp;500ml bag capacity);</li> <li>• Standard 15/22 mm Swivel connector allows connections to all common masks Endo-tracheal Tubes;</li> <li>• Provision to give supplemented oxygen-by-oxygen reservoir providing 100% oxygen;</li> <li>• Non-re breathing valve enabling the patient to inspire oxygen from the reservoir bag.</li> <li>• Should be single hand opera table</li> <li>• Should be easy to disassemble for cleaning and disinfection</li> <li>• Should have pressure release valve at 40cm H2O</li> <li>• Should have silicone oxygen tube 2m length.</li> <li>• 10. It should be upto 40 times autoclavable including bag and washers.</li> <li>• The bag should be made of soft silicone material.</li> <li>• Self-Inflating Resuscitator bag should*be of medical grade silicone rubber.</li> <li>• The reservoir should be a PVC bag of 600ml capacity 500ml bag capacity and 1000ml for 750ml bag capacity.</li> </ul> <p><b>1 Accessories (mandatory):-</b> Silicon bellows, Non Rebreathing Valve, 2 meter oxygen tube, Guedel</p>

Items	Specifications
	Airway sizes viz 0 and 1, Neonatal Mask of 3 sizes viz 00, 0 and 1 <b>STANDARDS AND SAFETY</b> ISO 13485; Manufacturer / supplier should have ISO certificate for quality standard. Should be USFDA / CE (From notified body) approved product or BIS certified should meet ISO 10651-4 standard requirement
<b>Lab Materials</b>	
Rapid Pregnancy Testing Kit	<b>Urine Pregnancy Test:</b> <b>Intended of Use:</b> One step hCG Serum/Urine Combo Rapi-Card rapid test for the qualitative detection of human chorionic gonadotropin (hCG) in serum and urine. <ul style="list-style-type: none"> <li>• Serum/Urine Combo Pregnancy Test Cassette is a rapid test that qualitatively detects the presence of hCG in serum and urine specimens at the sensitivity of 20mIU/mL. The test utilizes a combination of monoclonal and polyclonal antibodies to selectively detect elevated levels of hCG</li> <li>• Result should be produced with 1minute.</li> <li>• Accuracy:99%</li> <li>• Sensitivity:20mIU/mL</li> <li>• The test strips should have inbuilt quality control to achieve the above accuracy.</li> </ul> <b>Kit Configuration</b> <ul style="list-style-type: none"> <li>• Urine Pregnancy Test Rapid Card</li> <li>• Disposable pipette</li> <li>• Instructions for use</li> <li>• Storage condition 2-30 degree</li> </ul> <b>Quality Standards:</b> <ul style="list-style-type: none"> <li>• The manufacturer should be ISO 13485 certified.</li> <li>• The strips should be USFDA/CE (IVD) approved.</li> <li>• The strips should be DCGI approved.</li> </ul>
Dipsticks for urine test for protein and sugar	<b>URINE Complete rapid test reagent strips :</b> <ul style="list-style-type: none"> <li>• Urine Reagent Strips are for in vitro diagnostic use only.</li> <li>• Indications for urine test strips:               <ul style="list-style-type: none"> <li>• Screening for prevention</li> </ul> </li> </ul>

B

Items	Specifications
	<ul style="list-style-type: none"> <li>• Treatment monitoring</li> <li>• Patient self-testing</li> <li>• Urine Reagent Strips provide tests for the following parameters:               <ul style="list-style-type: none"> <li>• Glucose</li> <li>• Bilirubin</li> <li>• Ketone (Acetoacetic acid)</li> <li>• Specific Gravity</li> <li>• Blood</li> <li>• pH</li> <li>• Protein</li> <li>• Urobilinogen</li> <li>• Nitrite</li> <li>• Leukocytes</li> <li>• Ascorbic Acid in Urine.</li> </ul> </li> <li>• The Urine Reagent Strips should be packaged along with a drying agent in a plastic bottle with a cap to provide complete air tight.</li> <li>• Each strip should be stable and ready to use upon removal from the bottle.</li> <li>• The entire reagent strip should be disposable.</li> <li>• Results are obtained by direct comparison of the test strip with the color blocks printed on the bottle label.</li> <li>• All the reagent strips should be withstand at a room temperature between 15°-30°C (59°-86°F) and out of direct sunlight.</li> <li>• The minimum self-life of the urine strips should be 1year unopened and minimum 3months once it is opened.</li> <li>• The required controlled shall be provided along with the strip packet.</li> <li>• The strip pack sizes should be of 25/50/100 sizes.</li> </ul>

Items	Specifications
	<ul style="list-style-type: none"> <li>Urinalysis test strips types               <ul style="list-style-type: none"> <li>Ketones- Single test</li> <li>Glucose, Protein &amp; pH- Three parameter</li> <li>Glucose, Protein pH, Leukocytes, Nitrites, Ketones, Bilirubin, Blood, Urobilinogen, and Specific Gravity-10 parameter</li> <li>Leukocytes and Nitrite-Special parameter</li> </ul> </li> </ul> <p><b>Quality Standards:</b></p> <ul style="list-style-type: none"> <li>The manufacturer should be ISO 13485 certified.</li> <li>The strips should be USFDA/CE (IVD) approved.</li> <li>The strips should be DCGI approved.</li> </ul>
<b>Items required for providing Speciality services</b>	
<b>Oral Health</b>	
Interdental cleaning aids	Wooden/plastic Triangular Sticks Interproximal Brushes
Dental Probe	Single Ended with round handle.Made of SS.Handle of size 12mm, and Graduated probe.CPI TN/UNC15 typemade of stainless still 410 /420 grade.
Universal tooth extractor	Dental Forceps of standard size foradult use . Straight curvature with Serrated work surface area. Made of Stainless Steel 410 grade. Instrument handle is of Englishparten. The set shall be consists of 12 items of universal dimensions.
<b>ENT</b>	
Torch	LED focusing torch with chrome/steel ribbed body (medium size) with battery
Tongue Depressor	Wooden flat disposable toungedepressor. Box size:100/Box
Mouth Gag	Size-5". Finger ring instrument with grip & locking system.Blades of 1" long & 1/4" wide.
Mouth Mirror	Angular mirror-2 piecesTo provide indirect vision To retract lips, cheeks, and tongue To reflect light into the mouth.Accurate image from flat surface mirrors, image magnified with concave mirrors. Size of 14mm -22 mm of Plane size.StainlessSteel.Fog free with Flat/round ribbed handle. The mirror shall be made of fiber glass and the handle shall be made of stainless steel 410/420 grade
Tuning fork	Nickel-plated steel, material thickness 8 mm, Frequency: 128 Hz / 512 Hz
Nasal Speculum (St. Claire's)*	Nasal Speculum of length 3" having two 5mm wide x 11mm long blades.
Ear Speculum –	Ear Speculum – metallic, dull finish

Items	Specifications
metallic, dull finish	
Jobson-Horne probe	<p>Jobson-Horne probe-Curette loop at one end and threaded section at the other end for holding cotton wool            Manufactured from carbon filled nylon material to provide better strength and flexibility.</p>
Otoscope	<p><b>Otoscope</b></p> <ul style="list-style-type: none"> <li>• GMDN name: Otoscope</li> <li>• Clinical purpose: An otoscope or auriscope is a hand-held and battery powered device containing illumination and viewing optics medical device which is used to look into the ears. Health care providers use otoscopes to screen for illness during regular check-ups and also to investigate ear symptoms. An otoscope potentially gives a view of the ear canal and tympanic membrane, or eardrum.</li> <li>• Used by clinical department/ ward: ENT</li> <li>• Technical characteristics (specific to this type of device):           <ul style="list-style-type: none"> <li>• Battery (3.5v) operated high efficiency Fiber optic LED otoscope with detachable head and handle with high quality optics.</li> <li>• The viewing window with 3x magnification.</li> <li>• Should have on/off button on the handle for illumination, the handle should be made of Solid metal- chrome slip type shock proof.</li> <li>• The light should have minimum colour temperature of 4000k with CRI &gt;90 for Bright and homogeneous illumination with excellent colour rendering.</li> <li>• Should have rotating knob to control the intensity of the otoscope.</li> <li>• The LED lamp life should be more than 10000 hrs.</li> </ul> </li> <li>• User's interface: Manual</li> <li>• Software and/or standard of communication (where ever required) :NA</li> <li>• <b>PHYSICAL CHARACTERISTICS</b> <ul style="list-style-type: none"> <li>• Dimensions (metric): Hand Held Portable</li> <li>• Weight (lbs, kg): NA</li> <li>• Configuration : NA</li> <li>• Noise (in dBA) : NA</li> <li>• Heat dissipation: NA</li> <li>• Mobility, portability : Handheld device</li> </ul> </li> <li>• <b>ENERGY SOURCE</b> (electricity)           <ul style="list-style-type: none"> <li>• Power Requirements : NA</li> </ul> </li> </ul>

Items	Specifications
	<ul style="list-style-type: none"> <li>▪ Battery operated : Yes</li> <li>▪ Tolerance (to variations, shutdowns): NA</li> <li>▪ Protection : NA</li> <li>▪ Power consumption : NA</li> <li>• ACCESSORIES, SPARE PARTS, CONSUMABLES <ul style="list-style-type: none"> <li>▪ Accessories (mandatory, standard, optional); Spare parts (main ones); Consumables / reagents (open, closed system) <ul style="list-style-type: none"> <li>✓ Battery -2nos</li> <li>✓ Reusable EAR specula of 2mm, 3mm, and 4mm three from each. The specula should be autoclavable.</li> <li>✓ Storage case (rigid and steady)</li> </ul> </li> </ul> </li> <li>• ENVIRONMENTAL AND DEPARTMENTAL CONSIDERATIONS: <ul style="list-style-type: none"> <li>▪ Atmosphere / Ambiance (air conditioning, humidity, dust ...) <ul style="list-style-type: none"> <li>✓ Operating condition: Capable of operating continuously in ambient temperature of 10 to 40 deg C and relative humidity of 15 to 90% in ideal circumstances.</li> <li>✓ Storage condition: Capable of being stored continuously in ambient temperature of 0 to 50 deg C and relative humidity of 15 to 90%.</li> </ul> </li> <li>▪ User's care, Cleaning, Disinfection &amp; Sterility issues Disinfection: Parts of the Device that are designed to come into contact with the patient or the operator should either be capable of easy disinfection or be protected by a single use/disposable cover.</li> </ul> </li> <li>• STANDARDS AND SAFETY <ul style="list-style-type: none"> <li>▪ Certificates (pre-market, sanitary, ..); Performance and safety standards (specific to the device type); Local and/or international <ul style="list-style-type: none"> <li>✓ Product should be USFDA/CE approved</li> <li>✓ Should have IEC 60601-1/IEC 60601-1-2/CE (EU) certificate;</li> <li>✓ Manufacturer / supplier should have ISO 13485 certificate for quality standard;</li> </ul> </li> </ul> </li> <li>• TRAINING AND INSTALLATION <ul style="list-style-type: none"> <li>▪ Pre-installation requirements: nature, values, quality, tolerance: NA</li> <li>▪ Requirements for sign-off Certificate of calibration and inspection from the manufacturer</li> <li>▪ Training of staff (medical, paramedical, technicians)</li> </ul> </li> </ul>

Items	Specifications
	<ul style="list-style-type: none"> <li>✓ Training of users on operation and basic maintenance;</li> <li>✓ Advanced maintenance tasks required shall be documented</li> </ul> <ul style="list-style-type: none"> <li>• <b>WARRANTY AND MAINTENANCE</b> <ul style="list-style-type: none"> <li>• Warranty: 3 years including bulb</li> <li>• Maintenance tasks               <ul style="list-style-type: none"> <li>✓ Maintenance manual detailing;</li> <li>✓ Complete maintenance schedule;</li> </ul> </li> <li>• Service contract clauses, including prices:               <ul style="list-style-type: none"> <li>✓ The spare price list of all spares and accessories required for maintenance and repairs in future after guarantee / warranty period should be attached.</li> <li>✓ Free servicing (min. 2/year) during warranty period</li> </ul> </li> </ul> </li> <li>• <b>DOCUMENTATION</b> <ul style="list-style-type: none"> <li>• Operating manuals, service manuals, other manuals Should provide 2 sets(hardcopy) of:-               <ul style="list-style-type: none"> <li>✓ User, technical, maintenance and service manuals to be supplied along with machine diagrams;</li> <li>✓ List of equipment and procedures required for local calibration and routine maintenance;</li> <li>✓ Certificate of calibration and inspection;</li> </ul> </li> <li>• Other accompanying documents List of important spares and accessories, with their part numbers and cost;</li> </ul> </li> <li>• <b>NOTES</b> <ul style="list-style-type: none"> <li>• 11.1 Service Support Contact details (Hierarchy Wise; including a toll free/landline number);Contact details of manufacturer, supplier and local service agent to be provided;</li> <li>• Any Contract (AMC/CMC/add-hoc) to be declared by the manufacturer;</li> <li>• Recommendations or warnings: Any warning signs would be adequately displayed</li> </ul> </li> </ul>
<b>Communicable Diseases</b>	
Malaria Rapid Test Kit	Central store/NVBDCP
Whole blood finger prick HIV Rapid Test & STI screening test	<p><b>Intended of Use:</b> The assay should be able to detect antibodies of HIV1, HIV2 and all the subtypes by detection of antibodies by the agglutination/ Enzyme</p> <p><b>Should be 3rd generation:</b></p> <ul style="list-style-type: none"> <li>• The assay should have sensitivity of 100% or more and specificity</li> </ul>

B

Items	Specifications
	<p>of 100% or more as per data from an identified national reference laboratory.</p> <ul style="list-style-type: none"> <li>• The assay should have solid phase/ particles coated with synthetic and/ or recombination or both types of antigens of HIV1 &amp; HIV2.</li> <li>• Total procedure time should not be more than 30 minutes.</li> <li>• The manufacturers should ensure that: <ul style="list-style-type: none"> <li>▪ The test kit should be packed such that there is a provision to conduct single test at a time;</li> <li>▪ The assay components should include HIV positive and negative serum controls sufficient for conducting 20% of the tests (10% negative and 10% positive controls); and</li> <li>▪ The pack size of HIV rapid test kits should be 30 tests per Kit.</li> </ul> </li> <li>• The manufacturer should be ISO 13485 certified.</li> <li>• The strips should be USFDA/CE (IVD) approved.</li> <li>• The strips should be DCGI approved.</li> </ul>
Typhoid rapid test kit	<p><b>Widal test KIT</b>  <u>The test kit should have the following configuration</u></p> <ol style="list-style-type: none"> <li>1 'O' Antigen 5ml</li> <li>2 'H' Antigen 5ml</li> <li>3 AH' Antigen 5ml</li> <li>4 BH' Antigen 5ml</li> <li>5 Positive control 5ml</li> <li>6 Negative control 5ml</li> <li>7 Test Serum Sample 2 ml</li> <li>8 Glass Slide 1 No.RT</li> <li>9 Disposable Mixing Sticks</li> </ol> <ul style="list-style-type: none"> <li>➤ Result should be within 3 minutes</li> <li>➤ Homologues antigen antibody reaction with no cross reactivity with other salmonellar groups</li> <li>➤ High specificity:98%</li> <li>➤ Higher sensitivity:98%</li> <li>➤ Self-life 1 year</li> </ul>
Rapid test kit for	<b>Intended Of Use:</b> HBsAg/HCV Ab Rapid Test is a lateral flow



Items	Specifications
	<p>should work min. 3 months after opening of strips pack.</p> <ul style="list-style-type: none"> <li>• Control solution for checking reliability of strips will be supplied free of cost as &amp; when required.</li> <li>• Ready availability of reagent test strips, battery &amp; other consumables across Odisha for at least 5 years.</li> <li>• Machine should be supplied with lancing device of 2nos.</li> <li>• <b>Machine should have 4 yrs. of replacement warranty.</b></li> </ul> <p><b>Equipment Configuration:</b></p> <ol style="list-style-type: none"> <li>1-Glucometer-1no</li> <li>2-Lancing Device-2no</li> <li>3-Standard batteries-1Set</li> <li>4-Carrying case-1</li> <li>5-Instruction manual</li> <li>6-Warranty card</li> </ol> <p><b>Consumables:</b></p> <ol style="list-style-type: none"> <li>1. 115 nos. single use auto-disabled lancets in multi packs.</li> <li>2. Test strips -100 nos. in two packs</li> <li>3. Control strip.</li> </ol>
VIA Kit for screening cervical cancer	<p><b>VIA –Visual Inspection with acetic Acid kit:</b></p> <ul style="list-style-type: none"> <li>• Acetic Acid 0.5%=100ml=2bottles</li> <li>• Cotton swabs-1 pack of 100 swabs</li> <li>• LED focusing torch with chrome/steel ribbed body (medium size) with battery</li> <li>• Glass Slides -1 packet of 100 slides</li> <li>• Cosco speculum Made of SS 410 grade Medium Size</li> <li>• Sterilized disposable gloves-1 packet of 100 gloves</li> <li>• Good Quality bag/pouch for storing the above mentioned items</li> </ul>
<b>Ophthalmic services</b>	
Snellen vision chart	Snellen vision chart
Near vision chart	Near vision chart
Color Vision Chart	Color Vision Chart

**FW family planning Bio medical Equipment**

- 1.NSV KIT
- 2.IUCD Kit
- 3.Minitap Kit
4. PPIUCD

## 12. The lists of items of Physiotherapy & geriatric Equipment & Instrument

1. Examination Table
2. Wheel chair-
3. Partition screens-
4. Traction equipment: shoulder wheel and pulley or combo.
5. Height adjustment walker-
6. Height adjustment walking Stick-
7. BIPAP (Continuous positive air pressure)/ CPAP (Billable positive air pressure) -
8. Paraffin wax bath-
9. Pulse Oximeter finger-
10. Infrared radiation Therapy unit-
11. Patient Monitor (5 Parameter)-
12. Interferential Therapy unit(IFT)-
13. TENs (Transcutaneous Electrical Nerves stimulator)-
14. Nebulizer (Heavy duty)-
15. Peg board-
16. Hand therapy unit.
17. Quadriceps Table-
18. Weighing machine-
19. Head light
20. Foreign Body Removal Forceps
21. Tooth Model
22. Tooth Extraction Forceps set
23. Binocular Microscope
24. Stadiometer
25. Peak Flow Meter
26. Hydro Collator Unit
27. Dual frequency ultrasound therapy machine
28. Muscle stimulator (Electrical)

### TECHNICAL SPECIFICATION

#### **I. Lumbar & Cervical Traction with Bed**

1.1: Clinical Requirement: A cervical/lumbar traction device is used to help relieve pain associated with Cervical & Lumbar spine in case of Prolapsed inter vertebral disc, Listhesis of vertebra, Spinal nerve root Compression etc. and nerve root impingement. It also Spine and skeletal system, gently stretching the tight muscles and soft tissues of the spine area. By helping to open the space between the compressed vertebrae, a traction device assists in releasing the tension and pressure of the vertebrae structure.

#### **1.2 Technical Specification:**

- Microprocessor/ microcontroller based unit for treatment of cervical & lumbar regions of patient.
- It should be microprocessor controlled for smooth traction effect.

- The therapy time, HOLD and REST times should be easily programmable.
- Should be provided with Adjustable, single size PELVIC and THORACIC belts and the head halter fit for adult patient,
- The unit should have LED indicators for hold and rest time.
- Buzzer indicator for completion of treatment
- Separate selection for cervical and lumbar
- The unit should be sensitive to weight for accurate operation.
- Force – 4-90 Kg with doubler
- Cervical tension – 4, 6, 8, 10, 12, 14 (Kg)
- Lumbar tension – 18,21,24,27,30,33,40,45 (Kg)
- All the above increments are must be in steps.
- Mode: static and intermittent – cervical as well as lumbar,
- Hold time – 10-80 (Sec) adjustable in continuous mode.
- Rest time – 1-20 (Sec) adjustable in continuous mode
- Treatment time - 1 to 30 minutes –Adjustable.
- Patient control switch option should be there for safety.

### 1.3. Power Supply:

- The unit should work on 230volt & 50HZ supply
- The unit should have inbuilt over voltage protection.

### 1.4. Quality Standard:

- The model should be USFDA/CE/BIS certified
- The manufacturer should be ISO13485 certified
- The model should be compliance to electrical safety standards of IEC60601-1

### 1.5. TRACTION BED (MANUAL)

- Traction Bed – 3Fold sliding for horizontal cervical and lumbar traction
- Made of complete MS frame with wooden TOP
- Bed Dimensions: : 6' (L) x 2.5' (W) x 2.5' (H)
- Provided with appropriate dimension of mattress of 2inch thickness made of 40density PUF and cover with heat resistance rexin.
- Fully made on 30mm square tube with 18 gauge thickness.
- The lumbar section should roll freely on rollers for friction-free lumbar pull.
- The Cervical traction angle should controlled by vertical adjustment of the Machine mounting board.

**1.6. Flexion stool:** Adjustable Height range of 13"-20" which allows the therapist to place the lumbar spine into neutral or flexed position increasing patient comfort during traction treatments

The manufacturer should be ISO certified.

## 2. Interferential Current Therapy Unit (Combo)

**2.1: Clinical Requirement:** is an effective therapy option used for relieve of pain and accelerate the self-healing process, getting your body back to a healthy, pain Free State. The high frequency signals of an IFC penetrate through the skin into deeper

lying muscle tissues. It also improves circulation to muscles and reduces inflammation, stimulate muscles to contract.

#### 2.2: Technical Specification:

- Digital four channel with LED display interferential therapy device.
- Rotating Multi vector scale facility
- Pre-programmed memory with treatment applications
- Carrier frequency 2000 Hz and 4000 Hz
- Base frequency 0-150 Hz (1Hz step) Continuously adjustable
- Sweep frequency 0-150 Hz (1Hz step), Continuously adjustable
- Sweep modulation: 1/1 sec, 1/5/1/5 sec, 6/6/ sec.
- Therapy modes – two pole linear, four pole linear, four poles trapezoidal and four poles non-linear, four pole vector, Russian current and tens mode, faradic and galvanic, plane galvanic, Plane DC mode.
- Output current 0 -100 mA
- Timer 0-60min
- Indication for improper placement of electrodes for channel 1 and channel 2 separately.
- With attached trolley

#### 2.3: Power Supply:

- The unit should work on 230volt & 50HZ supply
- The unit should have inbuilt over voltage protection.

#### 2.4: Quality Standard:

- The model should be USFDA/CE(Notified)
- The manufacturer should be ISO13485 certified
- The model should be compliance to electrical safety standards of IEC60601-1

### 3. Transcutaneous electrical nerve stimulation (TENS)

**3.1: Clinical Requirement:** Used for Pain Relief and Treatment. TENS stands for (Transcutaneous Electrical Nerve Stimulation). Which are predominately used for nerve related pain conditions (acute and chronic conditions), muscle spasm. TENS machines works by sending stimulating pulses across the surface of the skin and along the nerves to close pain gates to reduce pain & spasm.

#### 3.2: Technical Specification:

- A micro controller based multi programmable transcutaneous electric nerve stimulator
- The unit should be a table top model.
- The unit should have dual Independent Channels
- The unit should have Adjustable Timer ranging from 0-90sec
- Adjustable Frequency and Pulse Duration parameters
- Therapy mode: Continuous, burst, linear, trapezoidal, triangular and non-linear
- Tens Frequency: 2 Hz to 150 Hz Adjustable
- Pulse Amplitude:0-80 mA

- Pulse width 50 msec – 300 msec, variable
- Therapy mode: Continuous, burst, linear, trapezoidal and no-linear
- Parameter selection: Manual and programmed.
- Treatment timer: Digital timer
- Output display: Display for CH1 & CH2.
- With attached trolley

### 3.3: Power Supply:

- The unit should work on 230volt & 50HZ supply
- The unit should have inbuilt over voltage protection.

### 3.4: Quality Standard:

- The model should be USFDA/CE/BIS certified
- The manufacturer should be ISO13485 certified
- The model should be compliance to electrical safety standards of IEC60601-1

## 4. Muscle stimulator unit (FC/GC)

**4.1: Clinical Requirement:** An electric muscle stimulator (often referred to as an EMS machine or unit) is an electronic device used to deliver electrical impulses to the body, in order to make the muscles contract and use in case of treatment of paresis and paralyse muscles. It also use as a primary diagnostic tool to know the innervated & de-nervated state of muscle.

### 4.2: Technical Specification:

- The device should produce Faradic and Galvanic types of current with variable Pulse width
- Should have provision for Galvanic & Faradic Curve
- The unit should have LCD display for therapy type and time along with all user settings infaces.
- Facility to plot SD Curve and Ionotophoresis
- Pulse duration, rest period, timer, intensity control Buzzer
- The unit should have treatment modes: Galvanic, Int.Galvanic, Faradic, & Surged Faradic.

### 4.3: Galvanic Mode:

- Therapeutic time adjustable from 30 – 300 mSec in (1 mSec step)
- Diagnostic time adjustable from -0.1, 0.3, 1, 3, 10, 30, 100, 300 mSec
- Frequency : 18 PPM to 3000 PPM
- Repetition Time : 300 mS to 1 mS

### 4.4: Faradic Frequency - 100 PPM

- Contraction time: 1 Sec to 9 Sec.
- Relaxation time: 1 Sec to 9 Sec.
- Digital display for diagnostic model
- Patient safety should be available with no output while intensity is high by switching on the machine.

### 4.5: Accessories:

- Velcro strap of 2" wide and 3 meter long
- Standard accessories: different electrodes small and large sizes and shapes (two of each size), active electrode of small and large size (two

of each size) Both Carbon & steel electrode, 2 different size pen electrode.

- With attached trolley

#### 4.6: Power Supply:

- The unit should work on 230 volt & 50HZ supply
- The unit should have inbuilt over voltage protection.

#### 4.7: Quality Standard:

- The model should be USFDA/ CE(Notified)
- The manufacturer should be ISO13485 certified
- The model should be compliance to electrical safety standards of IEC60601-1

### 5. Therapeutic Ultrasound Unit( Dual Frequency)

**5.1: Clinical purpose:** For treatment of painful conditions involving the musculoskeletal and neuromuscular structures, reduce inflammation of muscles, tendons, repair muscle, tendon, ligaments etc. It is also very use full for sports injury management, diabetic & leprosy foot ulcer healing etc.

#### 5.2: Technical Specification:

- It should be double frequency ultrasound machine.
- Modes of operation should be in continuous and pulsed
- It should have at least in 4 different settings for Pulse ratio.
- Maximum output 2.5-3 watts/Sq.cm.
- Output frequency 1 MHz and 3MHz
- Treatment time should be adjustable from 0-30 minutes
- Automatic alarm should be their on completion of the treatment.
- Digital LCD display to indicate the output in  $w/cm^2$  and other required services.
- All settings should be provided using touch panel.
- Both the transducer should be high quality water and impact resistant.
- The beam type :Divergent/Collimated
- Transducer: The Effective radiating area should be of 0.5 to 0.8mm for 1MHz and 0.4 to 0.6mm for 3MHz.
- The Beam Non uniformity Ratio (BNR) for both the transducer should be 4-6.
- With attached trolley

#### 5.3: Power Supply:

- The unit should work on 230volt & 50HZ supply
- The unit should have inbuilt over voltage protection.

#### 5.4: Quality Standard:

- The model should be USFDA/ CE(Notified )
- The manufacturer should be ISO13485 certified
- The model should be compliance to electrical safety standards of IEC60601-1

### 6. Shock wave Therapy unit:

**6.1: Clinical Use:** shock wave therapy primarily used in musculoskeletal disorders for the treatment of tendon pathies (proximal plantar fasciopathy, lateral elbow tendinopathy, calcific tendinopathy of the shoulder, and patellar tendinopathy, etc.) and bone defects (delayed- and non-union of bone fractures, avascular necrosis of femoral head, etc.)

**6.1: Technical Specification:**

It should be a table top model:

**6.2: Console:**

- The unit should be a ballistic radial shockwave therapy-system with electromagnetic generator as projectile accelerator
- User interface with facility for adjustment of parameter
- The unit should have LCD display for display of parameter during the therapy.
- The unit should have inbuilt preset programme of at least 7nos with minimum of
- The Unit should have facility for Adjustment of the parameters during the course of the therapy.
- The unit should provide non-invasive treatment for pain associated with musculoskeletal system
- Energy levels from 60 to 185 mJ (equivalent 1-5 bar)
- The energy should be adjustable in steps of 10 mJ
- Pulse frequencies from 1 to 20 Hz
- Burst mode for Trigger Point Treatment
- Positive shockwave counter
- The unit should not produce any negative shock.
- The unit should indicate the preset treatment protocols in the programme menu.

**6.3: Shockwave Applicator Hand Piece**

- The hand piece should be light in weight and ergonomic in design for user friendly operation.
- The hand piece should be made of anodized aluminium casing.
- It should have inbuilt fan for cooling.
- Minimum length of the hand piece should be 23-25 cm and a 5-6cm diameter (max.)
- Weight ca. 0.85 -0.90 kg (with cable)
- The Life of the shock generation should be minimum of 2,000,000 shocks

**6.4: ACCESSORIES:**

1. Hand-piece, complete with a 15 mm applicator
2. Holder for hand-piece
3. 25 mm a applicator
4. Silicon covers, 10 pcs
5. Endopuls lotion, bottle 250 ml
6. Foot switch

- 7. Carrier case
- 8. With attached trolley
- 6.5: Power Supply:
  - The unit should work on 230volt & 50HZ supply
  - The unit should have inbuilt over voltage protection.
- 6.7: Quality Standard:
  - The model should be USFDA/CE(Notified)
  - The manufacturer should be ISO13485 certified
  - The model should be compliance to electrical safety standards of IEC60601-1

## 7. Therapeutic Laser Unit (Physiotherapy)

**Clinical Use:** Laser Therapy Equipments use for treatment of the pain, trigger point, wound healing, Soft Tissue Injury Treatment etc effectively.

### 7.1: Technical Specification:

- Table top model consists of laser console, delivery system with accessories.
- Operating Frequency range 2Hz - 20kHz
- The unit should have facility to show the power in joules once the therapy time is set and Vis-a-vis.
- Timing range between 0-99 Minutes
- The unit should provide with two probes as single point and cluster type along with scanner.
- Probe -1:
  - Power output range: 100mW at 905nm
- Probe-2:
  - Power output of 10mW at 640nm
- Scanner:
  - Should have inbuilt predefined programme for easy/ default operation
  - The unit should have LCD display for parameter set & delivered etc
  - Should have inbuilt memory for patient database.
  - The unit should have two separate channel for connecting two probes for simultaneous operation.
  - The unit should be operated in continuous and pulse modes of operation.
  - Laser Dosage: 0.1-100 J/cm<sup>2</sup>
  - Therapy area 0.1-100 cm<sup>2</sup>
  - Laser class 3B

### 7.2: ACCESSORIES:

- 1) Hand-piece
- 2) Foot switch
- 3) Carrier case
- 4) With attached trolley

### 7.3: Power Supply:

- The unit should work on 230volt & 50HZ supply
- The unit should have inbuilt over voltage protection.

#### 7.4: Quality Standard:

- The model should be USFDA /CE (Notified) approved.
- The manufacturer should be ISO13485 certified
- The model should be compliance to electrical safety standards of IEC60601-1

### 8. Infra Red Therapy unit with stand

**8.1: Clinical Use:** An infrared lamp is the means to give superficial thermotherapy. Superficial thermo therapy is effective in pain and stiffness relieving, fasciitis in number of cases like back pain, cervical pain, frozen shoulder. It also use in facilitating healing of chronic ulcers.

#### 8.2: Technical Specification:

- Out put Power: 150 Watt
- Cord length: 3meter.
- Insulation: Class II (double isolation)
- Type of lamps: PAR 38 E, 150 W + prismatic rings for more focus.
- Make of lamp: Philips / Osram

#### 8.3: Power Supply:

- The unit should work on 230volt & 50HZ supply
- The unit should have inbuilt over voltage protection.

#### 8.4: Quality Standard:

- The model should be USFDA /CE(Notified )/BIS approved .
- The manufacturer should be ISO13485 certified
- The model should be compliance to electrical safety standards of IEC60601-1, Complies to IEC 60335

### 9. PhysioHydrocollator unit

**9.1: Clinical Use:** A moist heat modality use as a superficial to medium penetration thermotherapy modality. Clinically use for chronic pain management & joint stiffness cases.

#### 9.2: Technical Specification:

- Water tank made up of Stain less steel inner & outer Cabin ( 20-25 Gauge thickness)
- High grade thermally insulated
- Static/Movable base
- Top lid of stainless still, thermally insulated with Fibre handle.
- Dimensions: :L (33 cm) x W (42cm) x H(69cm)
- Tank thickness: 20-22 Gauge
- Tank capacity: Should be such so as to accommodate 8 Packs
- Temperature range: 50-90 degrees Celsius
- Thermostatic temperature control (50 - 90 degree C) - Auto cut-off
- Thermal cut-out temperature
- Temperature accuracy: +/- 10%

- Heat-up time to 90 degree Celsius: in 8 hrs.
- Cool-down time from 90 degrees Celsius: 3 hrs.

9.3: Power supply

- Mains power: 220/230 V, 50/60 Hz
- Power consumption: 1000-1500 W

9.4: Quality Standard:

- The model should be CE/BIS approved.
- The manufacturer should be ISO13485 certified
- Safety tests: Conforms to IEC 60601-1

Accessories must include:

- Hydrocollator Moist Heat Hot Pac : Standard Size-2, Over size-2, Cervical Standard size-2, Hydrocollator Knee or Shoulder Hot Pac-2
- Cotton Terry towel (Adult size): 8 Nos. to be supplied.

**10. Paraffin Wax Bath Unit:**

**10.1: Clinical use:** Paraffin wax treatments are used for the Symptomatic relief of pain and stiffness due to Arthritis, Bursitis, and Tendonitis, Muscle strains or Sports-Injuries. It is basically use as a superficially thermotherapy modality in physiotherapy.

10.2: Technical Specification:

- Double walled Construction with adequate insulation made of SS body.
- Wax tank made of 18-20 gauge Stainless steel & anodized Aluminium cover Mounted on 5cm dia. Four casters.
- Capacity – 25 Kg (Min.) to hold 20 Kg Wax
- Heater - 2000 watts
- Thermostatic temperature control (30 - 90 degree C) - Auto cut-off
- Thermostat - 30 to 110o C
- Perforated steel plate to cover heating element for safety of patient
- The scope of supply includes 20 Kg of Wax.

10.3: Power Supply:

- The unit should work on 230volt & 50HZ supply
- The unit should have inbuilt over voltage protection.

10.4: Quality Standard:

- The model should be CE/BIS approved.
- The manufacturer should be ISO13485 certified

**11. Short wave diathermy unit (500W) for Physiotherapy**

**11.1: Clinical Use:** It is a deep heating thermotherapy modality used for pain relief, joint stiffness reduction in number of cases like post fracture rehabilitation, Back & neck pain, Frozen Shoulder, Arthritis etc.

11.2: Technical Specification:

- High power medical diathermy transmitting oscillator valve and solid state rectifier
- RF power source Vacuum generating valve
- Power output: 500 watts
- Operation Frequency: 27.12 MHz
- Wave length: 10 to 12 meters
- Automatic timer, (timer 0 – 30min)
- Intensity control: 5 step smooth control with gradual increase
- The unit should have digital display.
- Cooling fan should be provided
- Two sets of cable and treatment Pads for Adult and Paediatric should be provided. (Pad Type Electrode)
- Disc electrode of two different sizes to be provided.

#### 11.3: Power Supply:

- The unit should work on 230volt & 50HZ supply
- The unit should have inbuilt over voltage protection.

#### 11.4: Quality Standard:

- The model should be CE/BIS approved.
- The manufacturer should be ISO13485 certified

### 12. Vacuum Therapy Unit

#### 12.1: Clinical Use:

- Decrease or relieve pain and inflammation from injury or chronic conditions.
- Relieve deep muscular issues and reduces muscle spasms.
- Release and soften scar tissue and loosen post-surgery adhesions.
- Lift and stretch soft tissue to separate fascia and restrictions.
- Increase range of movement and flexibility in joints.
- Promote healthy circulation and open the energy flow of the body.
- Move stagnation and enable normal lymphatic flow.
- Enhance any athletic training program to improve performance.

#### 12.2: Technical Specification:

- It should be a table top model of dual channel.
- The unit should generate continuous and pulsed outputs with variable pulse speed
- The vacuum pressure should be adjustable as per requirement.
- Vacuum: 100 – 500 mbar continuously adjustable
- Therapy forms: continuous or pulsed
- Pulsed vacuum: 2 – 6 secs. Adjustable
- No. of vacuum electrodes: Four of 60-70mm

- The unit should have digital display.
- All required accessories should be provided with the unit for smooth functioning. Glass or plastic cups and a vacuum pistol, bulb or machine to create suction on the body surface.
- With attached trolley

12.3: Power Supply:

- The unit should work on 230volt & 50HZ supply
- The unit should have inbuilt over voltage protection.

12.4: Quality Standard:

- The model should be USFDA /CE(notified) approved.
- The manufacturer should be ISO13485 certified

### 13. Knee and Hip Auto Passive mobilise

**13.1: Clinical Use:** The mobiliser used during the first phase of rehabilitation following a soft tissue surgical procedure or trauma, Post fracture management of Femur, Tibia-Fibula, Knee & hip replacement etc. . This rehabilitation procedure follow-up for control post-operative pain reduces inflammation, provide passive motion in a specific plane of movement, and protect the healing repair or tissue.

13.2: Technical Specification:

- The device should provide anatomical motion of the knee & hip for the adult and pediatric patient.
- Gear mechanism for continuous operation.
- Can be used for both left and right portion of the patient.
- Must be well cushioned for comfortable feel.
- Should have adjustable flexion angle up to 120 degree
- Provided with patients safety switch.
- Treatment timer : 0 to 60 min
- Features for Up & Down control limit button to control reverse direction movement
- Flexion delay : 5, 10, 15 sec
- Extension Delay : 5, 10, 15 sec

13.4: Power Supply:

- The unit should work on 230volt & 50HZ supply
- The unit should have inbuilt over voltage protection.

13.5: Quality Standard:

- The model should be USFDA /CE approved.
- The manufacturer should be ISO13485 certified
- The model should be compliance to electrical safety standards of IEC60601-1

### 14. Shoulder, Elbow and Wrist auto passive mobiliser

**14.1: Clinical Use:** The mobilise used during the first phase of rehabilitation following a soft tissue surgical procedure or trauma, Post fracture management of Humerus, Radius-Ulna, frozen shoulder etc. This rehabilitation procedure follow-up for control post-operative pain reduces inflammation, provide passive motion in a specific plane of movement, and protect the healing repair or tissue.

**14.2: Technical Specification:**

- The device should provide anatomical motion of the shoulder (Flexion-extension, Abduction-Adduction (vertical & Horizontal plane), Circumduction & Elbow (Flexion-extension), wrist combination of ulnar deviation, flexion (range - 0-80 deg) and Dorsey flexion plus radial deviation (range: 0-80 deg for the adult and pediatric patient.
- Gear mechanism for continuous operation.
- Can be used for both left and right portion of the patient.
- Must be well cushioned for comfortable feel.
- Should have adjustable for range of motion
- Provided with patients safety switch.
- Treatment timer : 0 to 60 min
- Features for Up & Down control limit button to control reverse direction movement
- Range of motion delay (For each movement pattern) : 5,10,15,30,45 sec
- Range of motion delay ( For each movement pattern: 5,10,15,30,45 sec

**14.4: Power Supply:**

- The unit should work on 230volt & 50HZ supply
- The unit should have inbuilt over voltage protection.

**14.5: Quality Standard:**

- The model should be USFDA /CE approved.
- The manufacturer should be ISO13485 certified
- The model should be compliance to electrical safety standards of IEC60601-1

**15. Parallel Bar**

**15.1: Clinical Use:** It use for horizontal surface gait training, Coordination training of lower limb in various neurological conditions like paralysis, GBS, Parkinson etc.

**15.2: Technical Specification**

- The parallel bar should be a height adjustable; plat form based one which should be very simple for easy accommodation of multiple patients of age group.
- The handrails should be in single piece, circular in design stainless steel ribbed pipe of minimum diameter 1 1/2".
- Both the end of the handrails should be fitted with bumpers.
- The Uprights should be made of heavy gauge square 1 1/2" steel tubes.

- All parts should be powder coated except the handrail. The handrail should single piece stainless steel pipe
- The uprights are to be fitted on a 1 1/2" thick hard plywood base water and termite resistance. Both end of this platform should be curved slope for obstruction free. The platform should be stain finish. Both the end of the platform should have two anti-slip threads for none skidding.
- Safety treads at both ends of platform should be provided for safety of the patient when take turn.
- Models make it safer for patient to turn

#### 15.3: Minimum dimension of the structure:

- Length of the bar: 10feet
- Width of the bar: 25" – 28"
- Height adjustment of the Bar: 26" – 39"
- The manufacturer should be ISO approved.

#### 16. Stair climbing training unit

**16.1: Clinical Use:** It use for vertical gait training, Coordination training of lower limb in various neurological conditions like paralysis, GBS, Parkinson etc.

#### 16.2: Technical Specification

- The stair case shall be a Small Up and Down type with 6step UP and 4 step down.
- The patient can ascend and descend without turning.
- The stairs are to be 10" of minimum depth.
- The total length of the stair case should be fitted with two full width rails
- It should bear a load capacity under normal for all category of patient.
- Can be easily sited to any space.
- Dimension:
- For down four steps: 10"L x 30" W x 6" H
- For UP four steps : 10"L x 30" W x 4" H
- The base platform: 24"L x 30"W x 24" H
- Complete unit shall be made of good quality Poly Wood of water and termite resistance
- The supplier should be ISO certified.

#### 17. Adult Trampoline

**17.1: Clinical use:** Use for coordination training of Parkinson, tabes dorsalis and vestibular rehabilitation.

#### 17.2: Technical Specification:

- Compact round trampoline, shape- round.
- Dimensions: diameter of the mat minimum 350cm
- The mat made of polypropylene.

- The perimeter should be padding to cover the springs and protect jumpers from a fall or finger pinching. The padding should be made of
- Anti-UV treatment on the safety pad, bed and net for optimal outdoor use
- Minimum lateral installation clearance
- Minimum 40mm galvanised steel frame.
- Maximum user weight 200kg
- Height of the mat above ground minimum 1feet
- 5-6 legs U shaped legs.
- 65-70 springs

#### 17.3: Quality specification

- The product should be CE approved.
- The manufacturer should be ISO certified.

### 18. Tilt Table (Electrical)

#### 18.1: Clinical use:

- Tilt Table is primarily use to transitionally bring a patient into a progressively upright standing position. They are used with bedridden, wheelchair bound patients or with people who are unable to support their own weight due to neurological impairment or injury.

#### 18.2: Technical specification:

- It should function like an electric operated table with vertical to horizontal tilt is achieved with an easy-to-turn.
- Crank is angled to provide leverage for maximum lifting efficiency.
- The table should have three positioning (0-90 degree) straps for security on raising the patient.
- Should have minimum 28" wide surface with an extra-wide footboard (20"W x 18"D).
- Minimum 2"-thick high-density (40) foam/Moulded foam padding.
- Made of heavy-duty steel frame with vinyl upholstery.
- Surface area: 28"W x 78"LL x 33"H (approx.). approximate
- The table should be operate on 230Volt, 50Hz AC electricity supply

#### Standard Quality

- The manufacturer should be ISO 13485 certified
- The model should be CE/BIS approved.

### 19. Manual therapy Bed unit

**19.1: Clinical Use:** Use for examination bed, Manual therapy bed for providing mobilisation, stretching, passive movement etc.

#### 19.2: Technical Specification:

- A 3 section, variable height plinth.
- With breathing slot at head end
- This plinth is built for sustained heavy use.
- Length : 74" Approx
- Width : 28" (Approx.)
- Height Adjustment : 21 inches to 37 inches approx
- Head rest up and down, Gas Spring
- Couch up and down Motorised
- The frame and the base shall be made of MS with powder coating.

## **20. Wage board:**

### **20.1: Clinical use:**

- For gross-motor activities as rolling tumbling, walking up or downhill or just lying down to read or study. Excellent for positioning severely disabled children.

### **20.2: Technical specification:**

- Size : 10" X 24" X 26" & 12" X 24" X 26"
- Made up of plywood of 19mm Thickness.
- It should be covered with a polyvinyl coating cover.
- Should not collect sweat, odour or dirt and should be easily wiped & cleaned with a damp cloth.
- Should have non sticky & non skidding surface.
- Should provide cover bag for safe keeping & handling.
- The manufacture should be ISO certified.

## **21. Adult Roller & Balance board (wobbles' Board)**

### **21.1: Clinical use**

- It should be useful for adult & pediatric.
- It should be used for rocking from side to side, spinning around and balancing where a patient can stand or sit.

### **21.2: Technical Specification:**

- The board made up off 19mm thick ply wood.
- The board should be provided with anti skid profile and four integrated handlebars.
- It should be covered with a polyvinyl coating cover.
- Should not collect sweat, odour or dirt and should be easily wiped & cleaned with a damp cloth.
- Dimension:
  - Adult roller: length - 30", Radius - 8" approx.
  - Balance board: approximate 4' (L) x 2' (W) across and approximately 7 inches height.
- The manufacturer should be ISO certified.

## 22. Therapeutic Mat.

### 22.1: Clinical use:

- Used for various mat exercises like stretching, active exercises to geriatric patients, paralysis and other neurologically impaired patients.

### 22.2: Technical Specification

- Made of good quality foam of minimum density of 40kg/cm<sup>3</sup>
- The thickness of the mattress should be 4"
- Dimension: 6.5ft (L) and 3 ft(W)
- It should be covered with a polyvinyl coating cover.
- Should not collect sweat, odour or dirt and should be easily wiped & cleaned with a damp cloth.
- Should have non sticky & non skidding surface.
- Should provide cover bag for safe keeping & handling.
- The manufacture should be ISO certified.

## 23. Shoulder wheel, over head pulley, Shoulder ladder (set)

### 23.1: Clinical use:

- Used for improving mobility & strength of shoulder girdle & rotator cuff muscles in case of frozen shoulder, paresis & paralysis cases.

#### Technical Specification:

#### A. Shoulder Wheel:

- It should be a wall mounted one.
- The wheel for use of adult and pediatric
- The motion arc can be adjustable from 10 to 38° inches by adjusting the handle.
- The wheel shall be mounted on a two chrome plated height adjustable (8" to 26") rails.
- The resistance can be varied by turning the resistance knob.
- The manufacturer should be ISO certified

#### B. Over Head Pulley:

- Heavy duty, medical grade, shoulders pulley exerciser for physical therapy use.
- Over the door metal bracket allows easy one hand setup.
- Units with a door strap require two hands, which is difficult with an injured shoulder.
- Easily adjustable cord length for any height and for use in both seating and standing position.
- The overhead pulley should have the provision of wall mounting.

#### C. Shoulder Ladder:

- Wood Finger/Shoulder Ladder
- 32 vertical finger steps
- Solid wood with heavy topcoats
- Pre-drilled mounting holes
- The manufacturer should be ISO certified.

## 24. Foot Bath Unit

**24.1: Clinical Use:** Use for diabetic foot .leprosy foot for debridement of dead skins and softening the skins in case of neuropathy.

**24.2: Technical specification:**

- Foot bath tub of size-L-12 X B-14 x D-6 1/2
- With top lid to cover when not in use.
- 3 function vibration massage bubble massage water tempering
- Attached infrared light spots
- Attached massage naps
- Removable splash protection
- Non Slip rubber feet
- Cord length -6 meter with auto winder
- Power : 220 Volt, 50 Amp & 60-100W

## 25. Quadriceps training unit including weight cuffs & dumbbells:

**25.1: Clinical Use:**

- Use for strengthening of lower limb and abdominal muscles in cases of paresis, paralysis, post surgical rehabilitation etc.

**25.2: Product Eligibility Criteria:**

- Should be CE /ISI approved product.
- Manufacturer should be ISO 9001 certified for quality standards.
- Manufacturer should have ISO 18001 certification for Occupational Health & Safety Assessment Series (OHSAS)

**25.3: Technical Specifications:**

- 2 folded top.
- Should be made up of rectangular CRC frame with epoxy powder coating.
- Should have facility of arm rest with height adjustment.
- Should have 2 inch cushion top seat with high quality rexin cover.
- Should have back rest adjustment with angle range of 5 to 90 degree with locking facility.
- The torque unit should have two lever arms with one adjustable weight and other one providing fixed contact with patient.
- Should have facility of changing the angle between two arms for providing maximum resistance at any point in the range.
- Should have height adjustment facility for torque unit with locking facility.
- Load range should be with maximum 100kg.
- Weight cuffs (6 Nos.) : ½ Kg - 3 Kg (1 each)
- Dumbbell (6 nos.) : ½ Kg - 5 Kg (1 each)

## 26. Treatment wooden beds with Mattress

**Technical Specification:**

- Foam rubber padded top,
- Size 72" x 24".

- upholstered with rexine & had adjustable back rest of 18" length which can be adjusted in height for desired position
- Two drawers is provided underneath the top.
- Top is mounted on sturdily built well polished hardwood legs & frames. Total height 31"
- 2 step wooden stairs should be provided with the bed.

## **27. Hand rehabilitation Therapy unit**

**27.1: Clinical use:** Used for exercise of fingers, intrinsic muscles, wrist flexor & extensor etc. In cases of paresis & paralysis of upper extremity.

### **27.2: Technical specification:**

- Hard-wood table with laminated top, size 25" x 35" x 30" high.
- Fitted with 6 pulleys in a steel frame. Leather loops and nylon cord passing through the pulleys with adjustable hanging weight underneath for Fingers Hand, Wrist and Forearm exercises.
- Especially for Metacarpo-phalangeal & Inter-phalangeal joints.
- Provided with Supinator/Pronator and Wrist-Circum-duction wheel for Wrist exercises.
- Comes with Five sets of weights, each set consists of 5 weights of one each of 100, 200, 300, 400 and 500 grms.
- Attached hand gym kit board on table top. Which have following features
  - Laminated exercise board with storage drawer.
  - Have several holes to fix Pegs and springs.
  - Comes with One finger Disc, One Roller, Three Spring loaded Knobs, 3 Aluminum pegs and 3 Sets of Springs of different tensions.
  - The unit permits active flexion and extension exercises of the finger and grip.
- Therapy Putty: FIVE packs of 450gm each in five different resistances.
- Medicine Ball Set: FIVE balls, 1, 2, 3, 4 & 5kg.

## **28. Thera bands /Elastic bands (Flat type) :**

**28.1: Clinical use:** Stretch-It a resistive treatment & exercise system with unlimited options.

### **28.1.2: Technical specification:**

- With each roll of 15cm wide x 5 meter long bands. Full Set of Five rolls come in following FIVE resistances for adult use.
  - (a) Light: Yellow.
  - (b) Medium-Light: Red.
  - (c) Medium-Heavy: Green.
  - (d) Heavy: Blue.
  - (e) Extra-Heavy: Black.

**28.2: Grip Exerciser:** Use for strengthening of hand grip.

- Chrome plated finished, metallic grip exerciser is fitted on a 75mm wide laminated base.
- Suitable for patients with very weak / infirm hand.
- Gives heavy resistance with six detachable springs.
- The upper bar is calibrated in centimeters and is provided with a needle to mark & record the daily progress.

### 30. Diagnostic equipments

#### 30.1. BP apparatus:

- Technique: Aortic
- Shock proof
- Latex free, Light weight ; easy to carry
- Range of BP monitoring : upto 300mm Hg with a accuracy of +3Hg
- The housing of the meter should be made of good quality thermoplastic/ corrosion proof aluminum alloy.
- The insufflations bulb should be made of good quality material and should allow rapid insufflations.
- The pressure release valve should permit precise release of pressure and also allow fast deflation.
- Micro filter must provide to protect air release valve and measuring system.
- Spoon-shaped grip made of stainless steel.
- The device should be shock resistant
- Should be supplied with a good quality carrying case
- Should be supplied with following reusable cuffs: Child, and Adult.
- The cuff should be latex free. The cuff surface should be easily cleanable by wash.
- The equipment should have comprehensive warranty for 3 years. The calibration should be for 3 years free.
- The instruments should be CE marked (certificate to be submitted in technical bid)
- Manufacturer should be ISO 13485 approved

#### 30.2. X ray-reader/X-ray view Box: use for X-ray viewing

- Single plate X-ray viewing screen
- LED lamps controlled by dimming ballasts
- Control luminance without flicker
- Uniform level of illumination across the entire front panel
- Sealed flush with the inside face of the operating theatre wall
- Spring loaded clips to secure the X-ray
- Minimum lamp hour: 30000hr
- Colour temperature should be of minimum: 4500K

#### 30.3. Gonio-meter: Use for measurement of joint range of motion.

- 3 different size Goniometer to accurately measure the axis and range of motion in a particular joint.
- Manufactured from clear toughened plastic for unrivalled durability.

- 12 Inch Goniometer featuring three different scales, 360° degrees, Metric and Imperial.
- Calibrated with the ISOM System.
- Manufacturer should be ISO certified.

**30.4. Hand grip & pinch Dynamometers:** Use for hand grip test and muscle strength measurement.

- Maximum reading remains until the unit is reset
- Five position handle and body
- Strength reading can be viewed as pounds or kilograms
- CE Certified
- Supplied in a protective carrying case
- The manufacturer should be ISO certified.

**30.5 Knee hammer:**

- Used to test joint reflexes or bone fractures in the joints
- The Taylor hammer features a 7.5 inch chrome handle
- Has a colored triangular rubber head
- The manufacturer should be ISO13485 approved

**30.6 Plumb unit:** Metallic plum with nylon thread

**30.7 Stadio meter:**

- The measuring rod can be dismantled into several pieces and can be set easily.
- The scale must be printed along the side of the measuring rod.
- Measuring range (Both in cm & inch) : 20-205 cm and 8 - 81".  
Graduation of measuring rod: 1mm / 8inch.
- The structure should be made of ABS plastic.
- The product should be CE certified (certificate to be submitted in technical bid).
- The manufacturer should be ISO certified
- Warranty : 1 Year

**30.8 Weighing scale:**

- Sturdy dial type mechanical platform weighing machine for adult and children.
- Zero adjustment facility should be there.
- Resolution : 50 gm
- Range of weighing: 0-120kg
- The manufacturer shall have the valid manufacturing license and should have model approval by the legal metrological Deptt. And the weighing scale must be stamped by the by legal metrological Deptt. In case of distributor, the bidder should have valid distributor and repair license from legal metrological Dept., Govt. of Odisha.
- ISO 9001 certified manufacturer (certificate to be submitted). Warranty: 1 Year

**30.9 Inch tape:**

- Spring controlled.
- The measuring tape should be
- The measuring tape made of woven fabric.
- It should be cleanable. Measurement range - 5 foot
- The tape should have both inch and centimeter scale.
- The calibration certificate should be provided.

**30.10 Skin fold measurement Calliper:** Use to measure skin fold thickness for body fat analysis.

Technical specification:

- Measurement Range: 0 to 70mm
- Material: ABS
- 1 Body Fat Caliper
- Measure Tape
- Measurement Chart

### 13. Waste Management Item

Sl.No	Item	Specification	Rate inclusive of all taxes
1	Color Coded Bins with Bio Hazard symbol	Size: 10 liter, 30 liter & 50 Liter, HDPE Material with Flip cover ISO Certified Colors Red, Yellow, Blue, and Black. Each bin is branded with water proof material with Bio Hazard symbol. Colour Pictorial Message in Odia and English printed with Govt. of Odisha, CDMO, Kendrapara supply	
2	White Puncture Proof container with lid with Bio Hazard Symbol	Size: 5 liter, Plastic with an imprint of text on the bins in Odia and English printed with Govt. of Odisha, CDMO Kendrapara supply,	
3	Electric needle Syringe Terminator	Body shock proof product with One year warranty	
4	Manual needle syringe Destroyer/Hub cutter	Plastic and stainless steel	
5	Hepatitis B Vaccine(Biological E	10ml vial dose (10dose)	
6	Tetanus Toxoid	10ml vial dose	
7	PPE	(Gun boot) Gloves Heavy duty rubber	

		gloves, Cloth face mass, Cloth head cap Reusable Rubber /plastic apron	
8	Trolley	Iron Body, 25mmX25mmX5mm angle ,25mm wheel bar ,16G GI sheet, Fibre wheel , 3'X2'x1.5' size	
9	Portable weighing Machine (Scale with sling )	With capacity up to 50 KG	
10	AD Syringe	0.5 ml	
11	AD Syringe	1 ml	
12	Sodium Hypo chloride 5ltr Jar	1% Solution	
13	Color Coded Bins with Bio Hazard symbol	30ltr,10ltr (Foot operated)	

#### 14. LIST OF PANCHAKARMA AYURVEDA MEDICINE

<i>A:- ARISHTAM</i>		
SL.NO.	PRODUCT NAME:-	Remarks
1	ABHAYAAARISHTAM	
2	ASHOKARISHTAM	
3	AMRUTHARISHTAM	
4	ASHWAGANDHAARISHTAM	
5	DASHMOOLARISHTAM	
6	JEERAKAADYARISHTAM	
7	KHADIRAAARISHTAM	
8	MRUTHASANJIEEVANI	
9	SAARASWATHAARISHTAM	
<i>B:- AASAVAM</i>		
SL.NO.	PRODUCT NAME:-	
1	CHANDANAASAVAM	
2	KUMAARVAASVAM	
3	PANCHAKOLAASAVAM	
4	LOHAASAVAM	
5	SHAARIBAADYAASAVAM	
6	PIPPALYAASAVAM	

C:- CHOORNAM	
SL.NO.	PRODUCT NAME:-
1	AVIPATHICHOORNAM
2	HINGUVACHADI CHOORNAM
3	KALLYAANAKAAVALEHAM
4	KOLAKULATTHAADI CHOORNAM
5	NAASIKA CHOORNAM
6	THAALEESAPATRADI CHOORNAM
7	VARAA CHOORNAM
8	VYSHWAANARA CHOORNAM
9	PUSHYAANUGA CHOORNAM
D:- GULIKA	
SL.NO.	PRODUCT NAME:-
1	AMRUTHA GULGULU GULIKA
2	CHANDRAPRABHA GULIKA
3	DHAANWANTHARAM GULIKA
4	GULGULU PANCHAPALAM GULIKA
5	KAANCHANAARA GULGULU GULIKA
6	KAANKAAYANAM GULIKA
7	KAISHORA GULGULU GULIKA
8	LAKSHA GULGULU
9	MAANASAMITRA VATAKAM
10	MAHAAYOGARAADA GULGULU GULIKA
11	MANDOORA VATAKAM
12	SHIVA GULIKA
13	THRIPHALA GULGULU GULIKA
E:- KASHAAYAM	
SL.NO.	PRODUCT NAME:-
1	AAMAVAATHA HARAM KASHAYAM
2	AARAGWADHAM KASHAAYAM
3	ASHTAVARGAM KASHAAYAM
4	BALAAJEERAKAADI KASHAAYAM
5	CHIRUVILWADI KASHAAYAM
6	DASHMOOLA KATUTHRAYAM KASHAYAM
7	DHANADANAYANAADI KASHAAYAM
8	GANDHARVAHASTTHADI KASHAAYAM

9	GULGULU THIKTHAKAM KASHAAYAM	
10	GULOOCHYADI KASHAAYAM	
11	KOKILAAKASHAM KASHAAYAM	
12	MAHAARAASNAADI KASHAAYAM	
13	MAHAAMANJJISHTAADI KASHAAYAM	
14	MUSALEE KHADIRADI KASHAAYAM	
15	PRASAARANVAADI KASHAAYAM	
16	RAASNAASAPTHAKAM KASHAAYAM	
17	RAASNEIRANDAADI KASHAYAM	
18	VARANADI KASHAAYAM	
19	SAHACHARAADI KASHAAYAM	
	F-: LEHYAM	
SL.NO.	PRODUCT NAME-:	
1	BRAHMA RASAAYANAM	
2	ASHWAGANDHADI LEHYAM	
3	GOMOOHRA HAREETHAKI LEHYAM	
4	HARIDRAA KAANDAM	
5	VYOSHAADI VATAKAM	
6	BRAHMI GHRUTHAM	
7	JEEVANTHYAADI GHRUTHAM	
8	MAHAAKALLYAANAKA GHRUTHAM	
9	MAHAAKALLYAANAKA GHRUTHAM	
10	MAHAATHIKTHAKA GHRUTHAM	
11	PANCHAGAVYA GHRUTHAM	
	G-: AVARTHI	
SL.NO.	PRODUCT NAME-:	
1	DHAANWANTHARAM 101 AVARTHI CAPSULE	
2	KSHEERABALA 101 AVARTHI CAPSULE	
	H-: THAILAM	
SL.NO.	PRODUCT NAME-:	
1	GANDHARVVAHASTHAADI THAILAM	
2	ANU THAILAM	
3	SAHACHARAADI THAILAM	
4	PINDA THAILAM	
5	DHAANWANTHARAM THAILAM	
6	KOTTAMCHUKKADI THAILAM	

*[Handwritten signature]*

7	KSHEERABALA THAILAM	
8	PRASAARANEE THAILAM	
	I-: LEPAM	
SL.NO.	PRODUCT NAME-:	
1	JAATHYAADI GHRUTHAM	
2	RASOTTHAMAADI LEPAM	
3	MAHAATHIKTHAKA LEPAM	

CHECK LISTPlease put  in the respective box

## DOCUMENTS submitted

Sl. No.	Details	Provided or not	
		Yes	No
1.	Cost of tender document Rs. 1000 /-(Rupees one thousand ) only		
2.	Earnest Money Deposit Rs 20000/-(Rupees twenty thousand only)(Not Applicable for SSI UNITS)		
3.	List of items being quoted with specification and strength		
4.	Duly attested Valid up-to-date Photocopy of Manufacturing License/Drug License N.B.-Information will be supplied by the bidder as per the format given in annexure No-II		
5.	Details of Manufacturing Unit / contact person for Liaisoning / local office in Odisha		
6.	Valid up-to-date WHO GMP/ GMP certificate N.B.- Information will be supplied by the bidder as per the format given in annexure No-II		
7.	Valid up-to-date ISO Certificate N.B.- Information will be supplied by the bidder as per the format given in annexure No-II		
8.	Attested photo copy for Purchase order of Lab. Materials, Consumables, Reagents, Drugs and Medical Consumables, Instrument & Equipment from Govt./Corporate/PSU Hospitals in India in support of marketing experience for last 3 years. (Marketing experience for supplier to retailer is not allowed)		
9.	The annual average turnover of the bidder should be Rs 1 Crore or more during the last three financial years certified by a Chartered Accountant. (Annexure VII)		
10.	Authorization letter of the Manufacturer if the bidder is an authorized distributor (Annexure VI) N.B.- Information will be supplied by the bidder as per the format given in annexure No-II		



(To be submitted in Cover A -Technical Bid)

## DETAILS OF THE BIDDER &amp; LOCAL CONTACT PERSON

	Corporate Office (The address in which the purchase orders and payment details will be communicated)	Local Contact Person / Branch Office / Zonal Office / if any, in Odisha.
Name & Full Address		
Telephone Nos., landline		
Mobile		
Fax		
E – Mail		

Signature of the Bidder:  
With seal

Date:

Official Seal:

(To be submitted in Cover A - Technical Bid)  
DECLARATION FORM

I / We ..... having My / our  
 .....office the at  
 .....do declare that I / We have carefully read all the  
 terms & conditions of tender of the Chief District Medical Officer, Kendrapara, Odisha for  
 the supply of drugs and medical consumables. The approved rate will remain valid for a  
 period of one year from the date of approval. I will abide with all the terms & conditions of  
 the Tender Reference no. \_\_\_\_\_

I/We do hereby declare I/We have not been de-recognized / black listed by any State Govt. /  
 Union Territory / Govt. of India / Govt. Organization / Govt. Health Institutions for supply of Not of  
 Standard Quality (NSQ) items / non-supply.

I/We agree that the Tender Inviting Authority can forfeit the Earnest Money Deposit and  
 blacklist me/us for a period of 2 years if any information furnished by us proved to be false at the  
 time of inspection / verification and not complying with the Tender terms & conditions.

I/We further declare that I/We possess valid manufacturing license / Drug License bearing  
 No. (s) ..... Valid upto ..... I / We  
 ..... do hereby declare that I / we will supply the  
 \_\_\_\_\_ as per the terms, conditions & specifications of the tender document.

Signature of the bidder :  
 Seal Date

Name & Address of the Firm:



(To be submitted in Cover A -Technical Bid)  
**MANUFACTURER'S AUTHORISATION FORMAT**

To  
The Chief District Medical & Public Health Officer,  
Kendrapara, Deptt. Of Health & Family Welfare  
Govt. of Odisha.

Ref: Tender No. \_\_\_\_\_ Dated \_\_\_\_\_ for supply of drugs and  
medical consumables.

Dear Sir,

We, \_\_\_\_\_ are the manufacturers of \_\_\_\_\_  
\_\_\_\_\_ having factories at \_\_\_\_\_

1. Messrs \_\_\_\_\_ (name and address of the agent) is our authorized agent for  
sale of \_\_\_\_\_ (name of items)
2. We confirm that Messrs. \_\_\_\_\_ (name of the above agent) is authorized to  
submit a tender, and enter into a contract with for the above goods manufactured by us.

Yours faithfully,

\_\_\_\_\_

(Signature with date, name and designation)

For and on behalf of Messrs \_\_\_\_\_  
(Name & address of the manufacturers)

Seal

Note:

1. This letter should be on the *letterhead* of the *manufacturer* and should be signed by a person  
having the power of attorney to legally bind the manufacturer.
2. Original letter shall be attached to the technical bid.

(To be submitted in Cover A - Technical Bid)

ANNEXURE - VI

(To be furnished in the letter head of the Auditor)

**ANNUAL TURN OVER STATEMENT**

The Annual Turnover for Drugs and consumables of  
M/s \_\_\_\_\_ who is a  
manufacturing unit / Distributor for the last \_\_\_\_\_ years are given below and certified that  
the statement is true and correct.

Sl.No.	Year	Turnover in Lakhs (Rs.) <sup>₹</sup>
1.	2018 - 2019	-
2.	2019 - 2020	-
3.	2020-2021	-

Average Annual Turnover (for the above three years) (Rs.) \_\_\_\_\_

Date:  
Place:

Signature of Auditor/ Chartered Accountant

(Name in Capital)

Seal

Membership No.-

Registration No. of Firm

**Note:**

- To be issued in the letter head of the Auditor.
- Separate certificates should be furnished for different manufacturer in case the bidder is quoting products of different manufacturers.

**PROFORMA FOR PERFORMANCE STATEMENT  
(For the period of last three years)-**

Name of Manufacturer

Sl.	Order placed by (Address of purchaser) (attach documentary proof)*	Order no. & Date	Value of Contract (Rs.)	Date of Completion		Reasons for delay if any
				As per contract	Actual	
1						
2						
3						
4						
5						
6						
7						
8						
9						
10						

Signature and seal of the Bidder

- The documentary proof will be copies of the purchase order (during the last 3 years) indicating Contract No. and date/
- \*\* The documentary proof will be certificate from the consignee/end user indicating Contact No. and date.

<u>Drug and Medical Consumables.</u>						
Sl. No.	Name of the Items	Specification	Unit / Pack	Rate / Unit (Inclusive of all Taxes)	Mfg By	REMARK
1	2	3	4	5		

Annexure - VIII (B)

LABORATORY REAGENTS & X-RAY ARTICLES

Sl. No.	Name of the Items	Specification	Unit / Pack	Rate / Unit (Inclusive of all Taxes)	Mfg By	REMARK
1	2	3	4	5		
	<u>X-Ray Articles</u>					

Annexure - VIII (C)

INSTRUMENT & EQUIPMENT

Sl. No.	Name of the Items	Specification	Unit / Pack	Rate / Unit (Inclusive of all Taxes)	Mfg By	REMARK
1	2	3	4	5		